

IHE Change Proposal

Tracking information:

IHE Domain	Devices (DEV)
Change Proposal ID:	CP-DEV-0xx
Change Proposal Status:	Draft
Date of last update:	2025-03-xx
Person assigned:	Kurt Elliason

Change Proposal Summary information:

Examples for Pillar/Rack/Slot	
Submitter's Name(s) and e-mail address(es):	Kurt Elliason kurt.elliason@icumed.com
Submission Date:	2025-03-xx
Integration Profile(s) affected:	IPEC/DEC
Actor(s) affected:	DOR
IHE Technical Framework or Supplement modified:	IHE_PCDTF Vol 2, dated 2024-11-04
Volume(s) and Section(s) affected:	M.1.2.3 Definition of Pillar/Rack/Slot topology
Rationale for Change:	

Section M.1.2.3 Definition of Pillar/Rack/Slot topology

On line 5702 add the following example OBX

OBX|2|ST|68549^MDC_ATTR_PILLAR_ASSEMBLY^MDC|1.0.0.2|2^3&2^12&8^pillar-orientation-vertical^pillar-direction-right^slot-rack-direction-up|||||F|||||109eee8b-1ceb-51d2-8496-6d6fd61eb6ab^^0012210000000000^EUI-64

On line 5708 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^2^3|||||F

On line 5711 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^^^7|||||F

On line 5722 add the following example OBX

OBX|2|ST|68549^MDC_ATTR_PILLAR_ASSEMBLY^MDC|1.0.0.2|1^1^4^pillar-orientation-vertical^pillar-direction-right^slot-rack-direction-down|||||F|||||109eee8b-1ceb-51d2-8496-6d6fd61eb6ab^^0012210000000000^EUI-64

On line 5708 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^1^2|||||F

On line 5711 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^^^2|||||F

Examples for Pillar/Rack/Slot

On line 5746 add the following example OBX

OBX|2|ST|68549^MDC_ATTR_PILLAR_ASSEMBLY^MDC|1.0.0.2|1^1^4^pillar-orientation-horizontal^pillar-direction-down^slot-rack-direction-right|||||F|||||109eee8b-1ceb-51d2-8496-6d6fd61eb6ab^^0012210000000000^EUI-64

On line 5755 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^2^3|||||F

On line 5758 add the following example OBX

OBX|9|ST|68550^MDC_ATTR_PUMP_PILLAR_DETAILS^MDC|1.1.0.24|2^^^7|||||F