

B.3 ERR – Error Segment, update usage of ERR-2

IHE Change Proposal

Tracking information:

IHE Domain	Devices (DEV)
Change Proposal ID:	CP-DEV-029
Change Proposal Status:	Submitted
Date of last update:	2026-05-11
Person assigned:	Kurt Elliason

Change Proposal Summary information:

B.3 ERR – Error Segment, update usage of ERR-2	
Submitter’s Name(s) and e-mail address(es):	Kurt Elliason kurt.elliason@gmail.com
Submission Date:	2026-05-11
Integration Profile(s) affected:	Point-of-Care Infusion Verification (PIV)
Actor(s) affected:	Infusion Order Programmer (IOP); Infusion Order Consumer (IOC)
IHE Technical Framework or Supplement modified:	IHE DEV TF Vol 2 (DEV TF-2), Rev. 10.0 Final Text, dated 2024-11-04
Volume(s) and Section(s) affected:	Vol 2: Appendix B.3 (ERR-2)
Rationale for Change: Add ERR-2 to table B.3-1 and add instructions that ERR-2 should be used to identify the source of the error for PIV message application acknowledgments.	

In Appendix B.3 ERR – Error Segment, (page 93), update the table and the following text:

Table B.3-1: ERR - Error segment

SEQ	LEN	DT	Usage	Card.	TBL#	Element name
1	493	ELD	B	[0..1]		Error Code and Location
3	705	CWE	R	[1..1]	0357	HL7 Error Code
4	2	ID	R	[1..1]	0516	Severity
5	705	CWE	RE	[0..1]	0533	Application Error Code
6	80	ST	C	0..1		Application Error Parameter

Notes: ERR-1 is included in HL7 v2.6 for backward compatibility only. Within the context of IHE PCD, this field shall not be used.

ERR-3 and ERR-4 are required by HL7 v2.6

Use of Error Segment in the PIV Profile. This list of error codes that can occur during the processing of a PCD-03 message is consolidated from participating pump vendors. The application

B.3 ERR – Error Segment, update usage of ERR-2

acknowledgment from the IOC should contain the Code and Text in ERR-5.1 and ERR-5.2 respectively. ERR-5.9 can also be used to contain additional text related to the error.