

Annexure-IV

Calibration Information Sheet for Baseline Noise, Characteristic Mass and Precision, Autosampler Linearity using Cr

Company Logo Here

XX PHARMACEUTICALS LIMITED

117 Adams Street, Brooklyn, NY 11201, USA

Calibration Information Sheet for Baseline Noise, Characteristic Mass and Precision, Autosampler Linearity using Cr

Test Results:

Test Parameter	Specification	Test Result	Pass/Fail	Comments
Chromium Baseline Standard Deviation	≤ 0.002			
Chromium Characteristic mass	2.2 pg – 3.8 pg			
Relative standar Deviation	$\leq 2.0 \%$			
Correlation Coefficient	≥ 0.999			