Annexure-II

Calibration Information Sheet for Thermostatic Column Compartment

Company Logo Here **XX PHARMACEUTICALS LIMITED** **117 Adams Street, Brooklyn, NY 11201, USA**

Calibration Information Sheet for Thermostatic Column Compartment

Instrument Name:

Instrument ID No.:

Location	Frequency
Instrumental lab	Once in every six month

Column Temperature Accuracy

Flow Rate : 1.0 ml/min				
Set Temperature	**Observe Temperature**	**Accuracy**	**Limit**	**Result**
80.0°C			**≤ 3°C**	
40.0°C				

Column Temperature Stability

Flow Rate : 1.0 ml/min							
Set Temperature	**Observe Temperature**	**Average**	**Low**	**High**	**Stability**	**Limit**	**Result**
80.0°C						**≤ 1°C**	
40.0°C							

Calibration Date :	**Next Calibration Date :**

Done By :	**Checked By :**
Name :	**Name :**
Designation :	**Designation :**
Date :	**Date :**

www.pharmaguideline.co.uk