

Pyranometer



Contact Details:

Student coordinator: Ravi Beniwal

Email: ravi.20mez0013@iitrpr.ac.in

Faculty in-charge: Dr. Himanshu Tyagi

DRF staff coordinator: me.drf@iitrpr.ac.in

Technical specifications (in brief):

Delta Ohm LP 471 PYRA 02.5 (Range 0-2000 W/m²)

(Spectral range: 300-2800 nm)

Application of the equipment/machine:

For measurement of solar irradiation

Limitations of the equipment/machine:

To be used as per the instructions

Consumables/sample preparation of the machine:

None