

**5) What about for  $d^{18}O$  analyses on bulk material or cellulose?**

For these analyses, the presence of water in the sample will affect the isotope measurement, so the samples should be carefully dried and stored in the desiccator prior to analysis. If a long time has elapsed between initial drying and encapsulation, it's best to re-dry the samples.