NDH Hospital, NTEC

Contour plus (BGA) Internal QC Record

Clinic / Unit / Ward: OT _ Machine Number of BGA: 1 / 2 / 3 / 4 Hi / Lo Control Solutions

Expiry Date: Opened on:

To be discarded on:

Perform QC testing at least once daily / before use, repeat under the following conditions:

- A new bottle / pack of strip is opened
- QC testing result is out of the acceptable range.
- Patient's CBG result is questionable.
- BGA is dropped.
- Accuracy of blood ketone result is questioned.
- Battery of BGA has been changed.

Please change a new form when changing a new BGA / control solution

Quality Check Operator			Test Strip		Acceptable Range				Corrective Action(s) / Remarks
Date	Time	Initials	Expiry Date	Lot Number	Control Solution Hi mmol/L	Within Range (Y / N)	Control Solution Lo mmol/L	Within Range (Y / N)	

Please keep this record for 3 years

Last modified Dec 2025 NDHOT

Checking Record for

Contour Plus
Blood Glucose Meter

Machine No.	EAM	Serial Number	Period of use
1	2568378	7887E2128904	July- Dec
2	2568380	7887E2128903	July - Dec
3	2568381	7887E2128902	Jan - Jun
4	2568379	7887E2128901	Jan - Jun