

IHE Change Proposal

Tracking information:

IHE Domain	Devices (DEV)
Change Proposal ID:	CP-DEV-005
Change Proposal Status:	Submitted
Date of last update:	20/Jan/2023
Person assigned:	(assigned by Domain Technical Committee)

Change Proposal Summary information:

Add "NA" abnormal flag for OBX-8 to PCD-09	
Submitter's Name(s) and e-mail address(es):	Andy Iverson andy.iverson@medtronic.com
Submission Date:	20/Jan/2023
Integration Profile(s) affected:	PCD-09 Implantable Device Cardiac Observation
Actor(s) affected:	Device observation reporter, device observation consumer
IHE Technical Framework or Supplement modified:	Patient Care Devices Technical Framework
Volume(s) and Section(s) affected:	Volume 2, 3.9.4.1.2.5-3
Rationale for Change: If a parameter is required but a vendor's device does not have that parameter the message must contain something. The proposal is that NA would be populated by the vendor in this situation.	

Add "NA" abnormal flag for OBX-8, see table 3.9.4.1.2.5-3

(Based on physician input from IDCO Interoperability discussion on 1-Aug-2018)