Brain Stealth protocol

Scan notes:

Include hard palate, tip of nose, ears, 1cm above top of head, and down to C1.

Axial T1 MPRAGE always run after contrast administration

Revised 11/12/2021
Charge as: MR brain w limited; or can be included in MR brain w wo contrast
Scanner: 1.5T or 3T
Coil: Head Coil/Flex Coil (flex coil may be needed if anterior head coil cannot clamp)

Plane	Weighting	Slice	Gap	FAT SAT	Scan Range
AXIAL	T1	1mm	0mm	None	Whole brain/parallel to AC/PC line. Include 1cm of air above scalp and down to C1
OPTIONAL RUN IF REQUESTED					
AXIAL	T2 spc	1mm	0mm	None	Whole brain/parallel to AC/PC line. Include 1cm of air above scalp and down to C1
AXIAL	FLAIR	1mm	0mm	None	Whole brain/parallel to AC/PC line. Include 1cm of air above scalp and down to C1

AXIAL Image Planning:

