



Bid Form for the Procurement of Goods

BID FORM

Procurement of Various Medical Equipment

Date: _____
 Project Identification No.: _____

To: DON EMILIO DEL VALLE MEMORIAL HOSPITAL
 Bood, Ubay, Bohol

Having examined the Philippine Bidding Documents (PBDs) including the Supplemental or Bid Bulletin Numbers *[insert numbers]*, the receipt of which is hereby duly acknowledged, we, the undersigned, offer to *[supply/deliver/perform]* *Procurement of Various Medical Equipment*;

Item No.	Description/General Specification	Brand/Manufacturer / Source	No. of tests or quantity per kit/bot/box/vol	Unit	Quantity	Unit Price	Total
1	BEDSIDE COMMODE, MEDICAL			unit	5		
	FOR INFIRMARY WARD (1), MEDICAL WARD (1) & ASU (1), ISOLATION WARD III (2)						
	Removable pail and armrests						
	4 rubber tips to prevent slipping						
	Comes with a plastic bowl with cover.						
	leatherette upholstery cover on top						
	Made of chrome-plated steel.						
	WARRANTY: 2 YEARS ON PARTS AND LABOR						
	Delivery Period: 60 days						
2	BP APPRATUS WITH WHEELS, WITH ADULT, PEDIA, NEONATE CUFF WITH TROLLEY			unit	2		

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	FOR ASU & ISOLATION WARD III						
	At least 5 leg mobile stand						
	Size: 11 adult						
	Flexi port reusable						
	2 tube cuff						
	Premium inflation bulb and valve						
	Coiled tube (4.0ft/1.2m)						
	BP cuff are anti-bacterial treated						
	Latex free, mercury free						
	Laser engraved dial free						
	Jeweled movement						
	with accessed dial free and high contrast and pointer						
	with built-in storage basket						
	with free stethoscope for adult						
	WARRANTY: 2 YEARS ON PARTS AND LABOR						
	DELIVERY PERIOD: 60 DAYS						
3	CARDIAC PATIENT MONITOR			unit	6		
	FOR ISOLATION WARD III (4) & ASU (2)						
	Multiparameter patient monitor						
	•Technical specification:						
	-at least 12.1 inch color touch screen, 1280x800						
	Compatible with HIS by HL7						
	Capable to connect to existing Cardiac Monitoring system						
	-Up to 2,400 hours capacity of the internal memory						
	Parameters : at least EKG, SPO2, NIBP, Pulse, RESP, TEMP						
	Optional Parameters: EtCO2, IBP						
	- With alarm on different parameters;						
	- The alarm system meets the requirements of IEC 60601-1-8						
	High-resolution colour touchscreen						
	with capacitive touchscreen technology						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	with precise SPO2 measurement with SIQ technology						
	with Lead-off detection feature						
	with Multi-gesture control						
	Simultaneous display of all parameters at a glance						
	For use of adults, children and newborns in emergency medical services						
	- Capable to have a simultaneous display of 8 to 12 waveforms						
	Integrated alarm system when measurements surpass predefined critical thresholds						
	- Capable that has Clinical Assistant Assessment for:						
	: sepsis screening and treatment tracking tools						
	: early signs of deterioration or potentially critical illness						
	: Glasgow Coma Scale						
	: Pacer View						
	: CCHD (Critical Congenital Heart Disease) Screening						
	DIMENSIONS:						
	- At least 308x282x185 mm						
	-Main unit weight: not more than 3.6 kg						
	BATTERY:						
	Rechargeable lithium-ion battery						
	- Battery operation time not less than 2 hours						
	MEASUREMENT RANGE AND ACCURACY (NIBP):						
	- Measures range: at least 30bpm-300bpm						
	TEMPERATURE SPECIFICATIONS:						
	MEASUREMENT AND ACCURACY: Measurement range 0°C- 50°C (32°F-122°F)						
	ECG SPECIFICATIONS: With common mode rejection, ST measurement -2.5Mv to +2.5Mv (-25.0 mm to +25.0 mm)						
	LEAD TYPE: 3-Lead; 5-Lead; 6-Lead, 12-Lead and Auto						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	OVERLOAD PROTECTION: Loaded with different input-circuit voltages of 1 V peak-to-valley, no damage						
	HEART RATE:						
	HR Alarm Range and Step: Alarm Limit:						
	- Upper limit:17bpm- 295 bpm						
	- Lower limit: 16 bpm- 290 bpm						
	- Extreme Tachy: 60 bpm- 300 bpm						
	- Extreme Brady: 15 bpm-120 bpm						
	RESP measurement range and accuracy:						
	- Range: 0 rpm-200 rpm						
	-Accuracy: 0rpm-120 rpm: ± 1 rpm; 121 rpm-200 rpm: ± 2 rpm						
	SPO2 SPECIFICATION:						
	Accuracy: - within range 70%- 100%: $\pm 2\%$ in Adu./Ped. Mode (non-motion), $\pm 3\%$ in Neo. mode (non-motion)						
	- with range 0%- 69%: not defined						
	PR SPECIFICATION:						
	Measurement range and accuracy (SPO2): - Measurement range: 20 bpm-300 bpm;						
	- Resolution: 1bpm;						
	- Measurement error: ± 2 bpm						
	ACCESSORIES:						
	- with ECG 3 leads						
	SPO2 finger- Adult- reusable						
	SPO2 finger- pediatric- reusable						
	TEMP sensor/reusable/ body surface						
	-Blood pressure cuff for adults, 25-35cm, reusable						
	-Blood pressure cuff for pediatrics, 18-26cm, reusable						
	- Thermal recorder printer with 2 rolls each						
	CONNECTIVITY/ CONNECTIVITY PORTS:						
	Equipotential conductor port						
	Network port						
	USB port of Monitoring system						
	VGA port						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	- Multi-function port that may be used as:						
	: as the nurse call port						
	: synchronization defibrillation output port to output defibrillation synchronization signals						
	: Analog output port to output analog signals						
	: wired or wireless connectivity						
	DOCUMENTATION:						
	Authorized distributoship certificates from The manufacturer under specific hospital account						
	WARRANTY: 2 YEARS ON PARTS AND LABOR						
	DELIVERY PERIOD: 60 DAYS						
4	ECG MACHINE, 3-CHANNEL			unit	4		
	FOR PACU, OB-ER, OPD & ISOLATION WARD III						
	"Technical Specification:						
	• Simultaneously 12 Leads Acquisition						
	• with at least 5 Inch HD LCD Screen with standard grid background						
	• weight less than 1kg						
	• Built-in print layout: at least 3x4, 3x4+1, 1x12						
	• External print /output layout: at least 3x4, 3x4+1,6x2 ,6x2+1,12x1						
	• with Sampling Modes: Real-time/ Pre-sampling/Triggered sampling/Periodic sampling						
	•with Manual/Auto/Rhythm Mode						
	•with Support Multiple Languages						
	•with Storage Format of Reports: ZQECG, PDF, JPG, PNG, XML, HL7, DICOM-WAVEFORM, DICOM-IMAGE, DICOM-PDF, GDT						
	•with Alarm: Electrode Disconnection, Recorder Unavailable, Tachycardia and Bradycardia and Low Power						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	•built in rechargeable lithium-ion battery						
	support USB and micro SD						
	Includes						
	Accessories: 2 roll ECG/thermal paper, Lead cable, Limb electrode, Chest electrode, Power cord, Earth wire, User manual						
	REquirement: Certificate of authorized distributor						
	ISO 13485						
	Warranty: 2 years warranty on parts & labor						
	DELIVERY PERIOD: 60 DAYS						
5	EMERGENCY CART MULTIFUNCTIONAL			unit	1		
	FOR ISOLATION WARD III						
	Technical Specs:						
	•At least 6 drawers						
	• With soft-close drawers glides						
	•Keyed or EMG breakaway central lock secures all drawers at the same time						
	•Push handle						
	•Whisper-quiet rolling convenience						
	•With hygienic wipe-down drawer						
	•ABS countertop surface						
	•With quality aluminum cart body construction						
	•ABS bumpers on each of the 4 corners						
	•Drawer Dimensions at least 22 in W x 16 in D / 55.9 cm W x 40.6 cm D						
	•Cart Dimensions atleast 27.8 in W x 23.0 in D x 42.3 in H (without guide handle) / 70.6 cm W x 58.4 cm D x 107.4 cm H (without guide handle)						
	Accessories:						
	• Side hanging box						
	• Waste bin						
	• Rotating tray / defibrillator shelf						
	• Telescopic IV pole						
	• Oxygen cylinder holder						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	• Cardiac board						
	• Two accessory rails						
	Warranty: 2 years warranty on parts & labor						
	DELIVERY PERIOD: 60 DAYS						
6	FETAL DOPPLER, HAND-HELD, with DISPLAY and CHARGER			unit	2		
	FOR OPD AND OB-ER						
	TECHNICAL SPECIFICATIONS						
	With dimensions of at least 168 mm × 31 mm × 67 mm						
	Weight must be 350 g or less						
	Probe cable length must be 2 m or longer and the weight must be 100 g or less						
	Display must be 0.96-inch OLED double color screen or better						
	The output power of the audio should be 2 W or better						
	FHR PERFORMANCE						
	The working mode must be continuous wave						
	The FHR measuring ranges must be from 50 bpm to 240 bpm and the resolution must be 1 bpm and accuracy must be ±2 bpm or better						
	The transducer parameters should have the following values						
	i. p - <1 MPa						
	ii. Iob < 20 mW/cm ²						
	iii. Ispta < 100 mW/cm ²						
	The effective radiating area of transducer must be 245 mm ² or better						
	The sensitivity should be 9 weeks gestation (3 MHz)						
	AUTO-SHUT DOWN						
	1						
	One (1) minute after no signal or operation, auto shut down						
	ACCESSORIES						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	3 MHz Probe						
	Carry Bag						
	WARRANTY: 2 years warranty on parts & labor						
	DELIVERY PERIOD: 60 DAYS						
7	HOSPITAL BED, MANUAL, 3 CRANKS, with SIDE RAILS & IV STAND, FOOT STOOL and BEDSIDE TABLE (BRANDED, HIGH QUALITY)			unit	5		
	FOR ISOLATION WARD III						
	1.General Specifications						
	a) At least three functions (Height adjustment, Back raise, knee raise)						
	b) With torque release handle for overwinding of shift protection.						
	c) Gas pressure spring height adjustment						
	d) 4 corners IV pole socket						
	e) With IV Pole storage under the bed, it can also be used as a Urine bag holder						
	2.Mattress base						
	a) Cold-rolled Steel Mattress Deck with Electrophoretic + powder double coating						
	b) With ventilation holes on the surface at least (5) per steel plate						
	c) The mattress deck can easily be raised for easy cleaning						
	3. Head Board /Foot Board						
	a) Removable head and footboard with lock						
	b) With hand grip design for both head and footboard						
	c) Board type Resin						
	4.Side Rail						
	a) Side-rail height at least 34cm						
	b) With lock indicator window. Green for locked and Red for unlocked						
	c) Collapsible side rails with at least 5 Pillars each side rail, aluminum tube material						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	d) Single switch gun handle structure design						
	e) Side rail mounting frame, the upper hinge zinc alloy die-casting material.						
	5. Technical Specifications						
	a) Overall length 217-218 cm						
	b) Total width 98-99 cm						
	c) Height Adjustment: 40-41cm Lowest, 73-74cm Highest						
	d) Mattress deck size at least L 200-201 cm x W 86-87 cm						
	e) Caster size at least 12.5 cm dia. Double-wheel with Central Locking System						
	f) Back rest of bed angle at least 0 - 75°						
	g) Knee angle adjustment at least 0 - 45°						
	6. Mattress:						
	a) At least 3-4cm thickness						
	b) Materials: Polyester cover, Urethane foam Padding						
	7. Inclusion:						
	a. IV Pole -one pc.						
	b. Bedside cabinet -one pc.						
	c. Overbed table -one pc.						
	d. Mattress – one pc.						
	e. alcohol-based rapid disinfectant with tuberculocidal effect,						
	active against enveloped viruses contains propan-2-ol, propan-1-ol and						
	N-(3-aminopropyl)-N-dodecylpropane-1, 3-diamine, certified VAH,						
	and efficacy based on DGHM methods -one liter						
	8. STANDARD DOCUMENTARY CERTIFICATIONS						
	a. Bidder's notarized certification that parts shall be available at the authorized Philippine service center/s for a period of 10 years after the warranty						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	b. Notarized certification that the brand of the equipment has been in the local market for the last 15 years						
	c. Certification from the manufacturer specializing in hospital beds for at least 15 years						
	d. Current and valid certificate of manufacturers compliance with ISO or its equivalent						
	e. Suppliers must provide certificate of installation from at least two (2) clients or more						
	f. Must have office in Cebu or Visayas						
	g. Bidders/Supplier’s Company must have track record of at least 5 years in existence specializing in medical device and equipment.						
	h. Hard Copies of product brochures/s or technical data sheet						
	i. Supplier must present Certification from previous clients that the hospital bed is in good						
	Working condition for at least 10 years.						
	Warranty: 2 years warranty on parts & labor						
	DELIVERY PERIOD: 60 DAYS						
8	INFUSION PUMP			unit	20		
	FOR ASU (5), Isolation ward III (5), DR (2), OB WARD (2), S-NICU (2), NICU (2), PEDIA (2)						
	Technical specifications:						
	Compact design,light						
	Stackable design						
	Connectable with workstation						
	Connectable with infusion central Monitoring system						
	Occlusion levels adjustable						
	has real-time Monitoring infusion states						
	has automatic anti-free-flow clamp						
	has Wi-Fi function available						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	Dimension: Not more than 202(W) x 74(H) x 133(D) mm						
	Weight not more than 1.22kg (including battery)						
	Has at least 2.4" LCD touchscreen, at least 7 colors selectable screen						
	Must be defibrillation-proof type CF applied part, IPX2						
	Compatible to all disposable IV sets (not brand specific) conforming to EN ISO 8536-4 standards						
	has handle						
	Can be directed at 90°/180°/270°/360° on pole clamp mounting						
	Must have infusion mode For Rate, time, weight, Sequence, Drip, Relay						
	KVO rate 0.1 – 5.0 ml/h adjustable						
	has power supply switching						
	Lithium battery 7.2 V, 2600 mAH , with battery duration ≥ 5 hours @ 25ml/h						
	Shall be compatible with procuring entity's existing infusion work station.						
	WARRANTY: 2 YEARS ON PARTS AND LABOR						
	DELIVERY PERIOD: 60 Days						
9	IV STAND W/ WHEELS, 4 HOOKS, STAINLESS			unit	50		
	FOR ISOLATION WARD II (10), ISOLATION WARD III (16), PEDIA (3), OB(3), NICU(2), S-NICU(2), PACU(3), INFIRMARY (3), OB-ER (3), MEDICAL WARD (2), SURGICAL WARD (3)						
	*at least 4 hooks						
	* stainless steel round tube and base						
	*Height can be adjusted from 135cm to 240cm						
	*Top with 4 chrome hooks, S.S.base with castors (Dia.50mm)						
	WARRANTY: 2 YEARS ON PARTS AND LABOR						
	DELIVERY PERIOD: 60 Days						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

10	LARYNGOSCOPE SET, ADULT			set	2		
	FOR ASU & ISOLATION WARD III						
	Fiber Optic Laryngoscope MacIntosh Set, with 3 blades, a handle, and Case						
	Specifications						
	Single-piece stainless steel construction ensures blade integrity						
	Comes with a removable light pipe and smooth surface design						
	Provides free repolishing service						
	Accessories includes:						
	Miller Laryngoscope Blade, Size 4						
	MacIntosh Laryngoscope Blade, Size 3						
	Lightweight "C" Size Handle						
	Lightweight "AA" Size Handle						
	Standard Case for MacIntosh Set						
	With extra halogen replacement bulb and free batteries for each handle						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
11	LARYNGOSCOPE SET, ADULT, PEDIA			set	2		
	FOR OB-ER & ER						
	Fiber Optic Laryngoscope MacIntosh Set, with 4 blades, 2 handles, and Case						
	Specifications						
	Single-piece stainless steel construction ensures blade integrity						
	Comes with a removable light pipe and smooth surface design						
	Provides free repolishing service						
	Accessories includes:						
	Miller Laryngoscope Blade, Size 1						
	Miller Laryngoscope Blade, Size 2						
	MacIntosh Laryngoscope Blade, Size 3						
	MacIntosh Laryngoscope Blade, Size 4						
	Lightweight "C" Size Handle						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	Lightweight "AA" Size Handle						
	Standard Case for MacIntosh Set						
	With extra halogen replacement bulb and free batteries for each handle						
	Blades must be compatible with procuring entity's existing blades						
	Warranty: 2 years warranty on parts and labor						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
12	LARYNGOSCOPE SET, PEDIA			set	2		
	FOR PEDIA WARD & NICU						
	Fiber Optic Laryngoscope MacIntosh Set, with 3 blades, a handle, and Case						
	Specifications						
	Single-piece stainless steel construction ensures blade integrity						
	Comes with a removable light pipe and smooth surface design						
	Provides free repolishing service						
	Accessories includes:						
	Miller Laryngoscope Blade, Size 1						
	Miller Laryngoscope Blade, Size 2						
	Lightweight "C" Size Handle						
	Lightweight "AA" Size Handle						
	Standard Case for MacIntosh Set						
	With extra halogen replacement bulb and free batteries for each handle						
	Warranty: 2 yearS service and parts						
	DELIVERY PERIOD: 60 DAYS						
13	NEBULIZER, MOTOR DRIVEN, HEAVY DUTY			unit	8		
	FOR ISOLATION WARD III (4), PEDIA WARD (2) & ASU (2)						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	Longest Duability Compressor With an Average Product Life of 10 Years						
	Metal Frame Motor Construction						
	Unique Four-Pole Motor Design Extends Life of the Unit						
	Large Built-In Compartment for Storing Nebulizers and Tubing						
	Built-In Nebulizer Holder						
	Includes Free Instructional DVD						
	Easy to Use						
	Product Specifications						
	Product Number: 5650D.						
	Width: 6.3 Inches						
	Height: 10.5 Inches						
	Depth: 10.1 Inches						
	Weight: 7.1 Pounds						
	Noise: 51 to 55 dBA						
	Maximun Liter Flow: 12 LPM						
	Maximum Air Pressure: 30 PSIG						
	Compressor Free Air Flow: 9 LPM or Greater						
	Product Life: 10 Years						
	HCPCS Code: E0570						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
14	PATIENT MOVER			unit	3		
	FOR ER, OR & DR						
	Features and Benefits:						
	Eliminates risk factors that often lead to patient or caregiver injuries.						
	Ensemble is covered with waterproof nylon for quick and easy cleaning.						
	Lightweight and rigid polypropylene material.						
	170 x 50cm, 2.5kg						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

15	PULSE OXIMETER, BEDSIDE CONTINUOUS BLOOD OXYGEN MONITORING SYSTEM			unit	2		
	FOR NICU & S-NICU						
	Technical Specification:						
	APPLICATION:						
	Neonatal, pediatric and adult patients						
	PHYSICAL DIMENSIONS & WEIGHT:						
	Base Unit: AT LEAST 8 x 4.5 x 4 (HxWxD inches)						
	Weight: at least 2.5 lbs						
	PERFORMANCE SPECIFICATIONS						
	Display : at least 5.0 inch (Diagonal) color TFT Resolution: at least 800 × 3(RGB) × 480						
	Trace: 2 waveforms						
	Waveforms: PLETH, ETCO2						
	Indicator: Alarm Indicator						
	Power indicator: Pulse beep and alarm sound						
	SPO2						
	ASpO2 : Anti-motion SpO2						
	SpO2% Range: 0 ~ 100%						
	SpO2 Accuracy: ±2% (70 ~ 100%,non-motion)						
	±3%(70 ~ 100% motion)						
	Pulse Rate Range: 30-250 bpm						
	Pulse Rate Accuracy: ±2 bpm(non-motion)						
	±3 bpm(motion)						
	Alarm Upper-lower Limit: Upper limit 70 ~ 100% Lower limit 70 ~ 100%						
	SpO2 Probe: Red light LED wavelength: 60nm±5nm						
	Infrared light LED wavelength: 40nm±10nm						
	Standards: Meets performance standards						
	EN ISO 9919:2005						
	INCLUSION:						
	Main unit						
	1 Pulse Oximeter						
	Standard Accessories						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	Adult Analog Finger Sensor For Ext. Cable						
	10 Pin Analog SPO2 Extension Cable						
	Analog Y Oximetry Sensor for Pedia/Neo						
	Multi Site Clip for Y Sensor						
	Power Cord (220 VAC)						
	Battery (5" Monitors Series)						
	Operators Manual CD						
	WARRANTY: 2 years warranty on parts and labor						
	DELIVERY PERIOD: 60 DAYS						
16	STETHOSCOPE, ADULT, HEAVY DUTY			unit	5		
	FOR ASU, ICU, ISOLATION WARD II, ISOLATION WARD III (2)						
	HEAVY DUTY						
	Brandnew						
	Adult						
	* Tunable diaphragms on either adult and pediatric						
	* single-piece diaphragm						
	* With big and small soft-sealing ear tips, non-cill bell sleeve, and directions						
	* No natural rubber latex or palate plasticisers						
	* Length: at least 69cm						
	* Weight: at least 150g						
	* special Adapter: yes						
	* Chest Piece						
	* Construction: Two-sided, adult and pediatric						
	* Material: Stainless Steel						
	* Diaphragm type: Tunable, single-piece						
	* Diaphragm Diameter: Adult: at least 4.3cm						
	* Construction: Single lumen (single acoustic channel within tube)						
	* Material: Not made with natural rubber latex or phthalate plasticisers						
	Headset						
	Materials: wide diameter aerospace alloy/anodized aluminum						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	EAR TIPS						
	Ear Tip type: Soft Sealing						
	Extra Ear Tips: yes						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
17	STETHOSCOPE, HEAVY DUTY (INFANT)			unit	4		
	FOR NICU, OB, DR & PICU						
	HEAVY DUTY						
	Brandnew						
	* Tunable diaphragms on either adult and pediatric						
	* single-piece diaphragm						
	* With big and small soft-sealing ear tips, non-cill bell sleeve, and directions						
	* No natural rubber latex or palate plasticisers						
	* Length: at least 69cm						
	* Weight: at least 150g						
	* special Adapter: yes						
	* Chest Piece						
	* Construction: Two-sided, adult and pediatric						
	* Material: Stainless Steel						
	* Diaphragm type: Tunable, single-piece						
	* Diaphragm Diameter: neonate: at least 3.3cm tubing						
	* Construction: Single lumen (single acoustic channel within tube)						
	* Material: Not made with natural rubber latex or phthalate plasticisers						
	Headset						
	Materials: wide diameter aerospace alloy/anodized aluminum						
	EAR TIPS						
	Ear Tip type: Soft Sealing						
	Extra Ear Tips: yes						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

18	STRETCHER SCALE			unit	1		
	FOR ER-ASU						
	FEATURES:						
	branded, portable/mobile						
	Can easily be lifted and rolled						
	accommodates all common stretcher types						
	with WIFI/Bluetooth connectivity						
	with AC adapter and battery						
	-equal or more than 450kg capacity						
	-platform size ± 10mm(1524mm(W) x 826mm(D)						
	-with 19 key, color-coded, sealed membrane keypad						
	-LCD display:(± 10) (89 mm W x 76 mm H) backlit blue						
	with at least 5 digit, 7 segment, high LCD weight digits						
	-ramp dimensions: (± 10) (10191 mm W x 826 mm D x 51 mm H)						
	-must have 2 or more integrated ramps						
	made of steel						
	-Keypad: 19 key, color-coded, sealed membrane						
	with hold/release key						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
19	SUCTION MACHINE, portable			unit	4		
	FOR ISOLATION WARD III (3) AND ASU (1)						
	Technical Specifications:						
	- Powerful-adjustable Vacuum power at -75 kPa / -563mmHg						
	- 18 Liters per minute flow rate						
	- Noise level at 51.2 dB(A) (Max. Operation)						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	- 4 modules Piston Cylinder Technology-Suction System						
	- With overflow/virus filter						
	- Vacuum regulator: Membrane type with safety knob (with locking mechanism)						
	- Battery operated for 30 minutes						
	- LED warning light with acoustic sound						
	- Vacuum regulator with safety knob						
	-Flexibility on the collection system						
	- Can be used with Disposable Liners System or Reusable Collection System						
	- IEC 60601 safety rating - Type CF (Cardiac Floating: Max earth leaking current						
	10µA) for protection against electric shock and direct conductive contact with heart						
	- Ingress protection: IP21						
	- Manufacturer Certification: CE / ISO approval						
	- Carrying handle for easy mobility						
	-With tubing holder						
	- 2 integrated front and side jar holder "						
	Inclusions:						
	- 1 unit AC/DC						
	-1 Overflow protection / virus filter set						
	- 1 Silicone tubing, Ø D7 x 12mm, 25cm with 2 coupling pieces (1pc 90° angle)						
	- 1 Suction Jar, Polycarbonate, 1.5L						
	- 1 unit Adapter						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						
20	SYRINGE PUMP			unit	5		
	FOR PACU (2), ISOLATION WARD III (2) & ASU						
	Technical specifications:						
	Compact, light and easy carrying						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

Stackable, Connectable with pumps or workstation						
Wi-Fi function available						
connectivity even in transport						
Compatible with HIS/CIS system						
Connectable with central Monitoring system						
Occlusion levels adjustable						
has real-time Monitoring infusion states						
-Dimension: Not more than 252(W) x 74(H) x 133(D) mm						
-Weight < 1.6kg (including battery)						
-Has at least 2.4" LCD touchscreen, with at least 7 colors selectable screen						
Must be defibrillation-proof type CF applied part, IPX2						
Compatible to all disposable syringes 10ml, 20, 30ml, 50/60ml IV sets (not brand specific) conforming standards						
infusion mode: Rate, time, weight, Sequence, Relay mode						
-Infusion rate range 0.1-300.0 ml/h (10ml syringe)						
0.1-900.0ml/h (30ml syringe)						
0.1-600.0ml/h (20ml syringe)						
-0.1-2000ml/h (50/60ml syringe)						
-KVO rate 0.1 – 5.0 ml/h adjustable, step by 0.01 ml/h, default rate 1ml/h						
has Relay function available with workstation						
-Event recording maximum 2000 events can be stored						
has power supply switching when AC/DC power supply is cut off						
-Battery duration ≥ 6 hours @ 5ml/h						
-Lithium battery 7.2 V, 2600 mAH						
has handle						
-Can be directed at 90°/180°/270°/360° on pole clamp mounting						
Includes: 1 (ONE) infusion workstations as free goods						
with technical specifications:						
-Power consumption: less than 100VA						

Printed Name and Signature of Proprietor/Manager/Authorized Representative

	-Syringe/infusion pump combination channel: 1-9						
	Must be Class I, IPX2						
	-Cabinet body not more than 254(w)x254(h)x152(d)mm						
	-Cabinet body weight less than 1.6kg						
	Warranty: 2 year warranty on parts and services						
	DELIVERY PERIOD: 60 DAYS						
21	WHEELCHAIR WITH IV STAND AND OXYGEN HOLDER			unit	5		
	FOR ISOLATION WARD III (3) & OPD (2)						
	Brand new, Heavy duty						
	- Adult size						
	- Foldable						
	- Chromed Steel Frame						
	- Fixed and padded Armrest						
	- Weight capacity: not less than 250 lbs.						
	- Seat widths: not smaller than 18"						
	- Seat depths: not higher than 16"						
	- With solid rubber spoke wheels on each side						
	- with Push/Pull lock System on every side.						
	- with two (2) solid casters in front.						
	- Fixed Footrest						
	- Double Cross Bar						
	with oxygen holder can fit the portable oxygen tank of 40 cm diameter						
	with IV stand 4 poles						
	Warranty: 2 years service and parts						
	DELIVERY PERIOD: 60 DAYS						

Grand Total _____

Printed Name and Signature of Proprietor/Manager/Authorized Representative

in conformity with the said PBDs for the sum of [total Bid amount in words and figures] or the total calculated bid price, as evaluated and corrected for computational errors, and other bid modifications in accordance with the Price Schedules attached herewith and made part of this Bid. The total bid price includes the cost of all taxes, such as, but not limited to: [specify the applicable taxes, e.g. (i) value added tax (VAT), (ii) income tax, (iii) local taxes, and (iv) other fiscal levies and duties], which are itemized herein or in the Price Schedules,

If our Bid is accepted, we undertake:

- a. to deliver the goods in accordance with the delivery schedule specified in the Schedule of Requirements of the Philippine Bidding Documents (PBDs);
- b. to provide a performance security in the form, amounts, and within the times prescribed in the PBDs;
- c. to abide by the Bid Validity Period specified in the PBDs and it shall remain binding upon us at any time before the expiration of that period.

[Insert this paragraph if Foreign-Assisted Project with the Development Partner:

Commissions or gratuities, if any, paid or to be paid by us to agents relating to this Bid, and to contract execution if we are awarded the contract, are listed below:

Name and address Amount and Purpose of
of agent Currency Commission or gratuity

(if none, state "None")]

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

We certify/confirm that we comply with the eligibility requirements pursuant to the PBDs.

The undersigned is authorized to submit the bid on behalf of [name of the bidder] as evidenced by the attached [state the written authority].

We acknowledge that failure to sign each and every page of this Bid Form, including the attached Schedule of Prices, shall be a ground for the rejection of our bid.

Name: _____

Legal capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Date: _____

Printed Name and Signature of Proprietor/Manager/Authorized Representative