

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
1	Airway	0 size, Rate / No	Ibrahim Enterprises UNICORN HEALTH	Tahir Pharma Medicare					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted					
2	Airway	1 size, Rate / No	Ibrahim Enterprises UNICORN HEALTH	Tahir Pharma Medicare					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality Rough surface	Accepted					
3	Airway	2 size, Rate / No	Ibrahim Enterprises UNICORN HEALTH	Tahir Pharma Medicare					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality Rough surface	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
4	Airway	4 size, Rate / No	Ibrahim Enterprises UNICORN HEALTH	J.J. Brothers Enterprises Well Lead	Tahir Pharma Medicare				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality Rough surface	Accepted	Accepted				
5	Airway	5 size, Rate / No	Ibrahim Enterprises UNICORN HEALTH	Tahir Pharma Medicare					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality Rough surface	Accepted					
6	Soda Lime	Pack of 4.5 Kg / 5 Litre, Rate/No.	Ibrahim Enterprises ATRASORB	Tanveer Enterprises Emedical Soda Lime	Tahir Pharma Kngmed				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
7	C.V.P Lines	All sizes (Single Lumen) Rate/No.	Friends Enterprises Fornia	Flowtronix Systems Able	Popular International Pvt. Ltd GM KEY Giant Medical Ltd China	Tanveer Enterprises ARROW			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Accepted			
8	Epidural Anaesthesia Set	All Size Rate/No	Friends Enterprises PORTEX	Noor International Akus	Husnain Pharma Perifix				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Filter absorbs drug More resistance when injecting	Accepted				
9	Instrument Disinfectant Solution of appropriate composition	Rate per bottle, Rate calculated after Dillution and per day cost will be calculated Pack of 5 liter or less	Friends Enterprises HICLEAN	Tanveer Enterprises Detro Cid Enzym	Tahir Pharma Triacid-N				
	Product Evaluation by TEC / End user.		Accepted	Accepted	Accepted				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
10	Endotracheal Tube 2.5 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Conector is loose Kinks easily	Rejected Not according to Specification Provided 7.5 instead of 2.5	Accepted	Sample not provided	Rejected Connector is very loose	Accepted	
11	Endotracheal Tube 3.0 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Not according to Specification Provided 7.5 instead of 3.0	Accepted	Sample not provided	Rejected Connector is very loose	Accepted	

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
12	Endotracheal Tube 3.5 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Not according to Specification Provided 7.5 instead of 3.5	Accepted	Sample not provided	Accepted	Accepted	
13	Endotracheal Tube 4.0 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Not according to Specification Provided 7.5 instead of 4.0	Accepted	Sample not provided	Accepted	Accepted	
14	Endotracheal Tube 4.5 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Connector is loose	Rejected Not according to Specification Provided 7.5 instead of 4.5	Accepted	Sample not provided	Rejected Connector is loose	Accepted	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
15	Endotracheal Tube 5.0 No.	(Disposable) Without cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Soft tube Loose connector	Rejected Not according to Specification Provided 7.5 instead of 5.0	Rejected Not according to Specification Provided 7.5 instead of 5.0	Sample not provided	Accepted	Accepted	
16	Endotracheal Tube 5.0 No.	(Disposable) With cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Soft tube Kinks at body temp. Loose connector	Rejected Not according to Specification Provided 7.5 instead of 5.0	Accepted	Sample not provided	Accepted	Accepted	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
17	Endotracheal Tube 5.5 No.	(Disposable) Without & with cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily Loose connector	Rejected Not according to Specification Provided 7.5 instead of 5.5	Accepted	Sample not provided	Accepted	Accepted	
18	Endotracheal Tube 6.0 No.	(Disposable) With cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily Loose connector	Rejected Not according to Specification Provided 7.5 instead of 6.0	Accepted	Sample not provided	Accepted	Accepted	
19	Endotracheal Tube 6.5 No.	(Disposable) With cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily Loose connector	Rejected Not according to Specification Provided 7.5 instead of 6.5	Accepted	Sample not provided	Accepted	Accepted	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
20	Endotracheal Tube 7.0 No.	(Disposable) With cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily Loose connector	Rejected Not according to Specification Provided 7.5 instead of 7.0	Accepted	Sample not provided	Sample not provided	Accepted	
21	Endotracheal Tube 7.5 No.	(Disposable) With cuff Rate/No	Friends Enterprises K&K Zenith	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily Loose connector	Cuff and ballon is of poor quality	Accepted	Sample not provided	Accepted	Accepted	
22	Endotracheal Tube 8.0 No.	(Disposable) With cuff Rate/No	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd SHILEY / Curity	Husnain Pharma Endosoft			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Sample not provided	Sample not provided	Sample not provided			

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
23	Nasogastric Tube No. 4	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Feeding Tube	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Poor material Short length	Rejected Poor material	Accepted	Accepted	Accepted	
24	Nasogastric Tube No. 5	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Feeding Tube	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Poor material Short length	Rejected Poor material	Accepted	Accepted	Accepted	

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
25	Nasogastric Tube No. 6	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Feeding Tube	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Thin material Kinks easily	Rejected Thin material Kinks easily	Rejected Thin material Kinks easily	Accepted	Accepted	Accepted	
26	Nasogastric Tube No. 8	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Feeding Tube	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Thin material Kinks easily	Rejected Thin material Kinks easily	Rejected Thin material Kinks easily	Accepted	Sample not provided	Accepted	

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
27	Nasogastric Tube No. 10	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Feeding Tube	J.J. Brothers Enterprises Well Lead	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Hard material Kinks easily	Rejected Hard material Kinks easily	Accepted	Accepted	Accepted	
28	Nasogastric Tube No. 12	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Nasogastric Tube	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Hard material Kinks easily	Rejected Hard material Kinks easily	Accepted	Accepted		

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
29	Nasogastric Tube No. 14	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Nasogastric Tube Husnain Pharma Nasogastric Tube	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Thin material Kinks very easily Resulting blockage	Rejected Thin material Difficult to maneuver Kinks easily	Accepted	Accepted		
30	Nasogastric Tube No. 16	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Nasogastric Tube	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Thin material Difficult to maneuver Kinks easily No radio-opaque line	Thin material Difficult to maneuver Kinks easily No radio-opaque line	Thin material Difficult to maneuver Kinks easily Not according to Specification provided 14Fr instead of 16	Accepted	Accepted		

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
31	Nasogastric Tube No. 18	Sterile Rate /No	Friends Enterprises PIGEON	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Nasogastric Tube	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Thin material Kinks easily	Sample not provided	Accepted	Accepted		
32	Nelton Drain No. 08	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Hard material	Sample not provided	Accepted	Rejected Hard material			
33	Nelton Drain No. 10	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Hard material	Sample not provided	Rejected Hard material			

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
34	Nelton Drain No. 12	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Rejected Poor quality Kinks easily	Accepted			
35	Nelton Drain No. 16	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Thin material Kinks easily	Sample not provided	Accepted	Accepted			
36	Nelton Drain No. 18	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Thin material Difficult to maneuver Kinks easily	Thin material Difficult to maneuver Kinks easily	Accepted	Accepted			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
37	Nelton Drain No. 20	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Thin material Kinks easily	Sample not provided	Accepted	Accepted			
38	Nelton Drain No. 22	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Sample not provided	Accepted	Accepted			
39	Nelton Drain No. 24	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Husnain Pharma Nelton Catheter	J.J. Brothers Enterprises Well Lead			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Thin material Difficult to maneuver Kinks easily	Accepted	Accepted			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
40	Nelton Drain No. 26	Sterile Rate /No	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Sample not provided					
41	Sterilized Surgical Blades All sizes	UK/USA/Japan/EU origin only Rate/No	Friends Enterprises FEATHER						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
42	Skin Grafting Blade (Sterilized)	Rate / No.	Friends Enterprises FEATHER						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
43	Urine Drainage Bag (adult)	2000ml Rate/No	Friends Enterprises Star	Ibrahim Enterprises CLASSIC	Pioneer Pharma Relation	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Ultra for Medical Company Egypt	Husnain Pharma Curaflow	J.J. Brothers Enterprises Well Lead
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor material Poor outlet	Accepted	Rejected Stopper not functional properly	Accepted	Sample not provided	Accepted	Accepted
			Tanveer Enterprises GB Urine Bag	Tahir Pharma Medicare					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
44	Urine Bag (Paeds)	Rate/No	Ibrahim Enterprises MEDILINE	Zainabain Enterprises Faisalabad WELLMED	J.J. Brothers Enterprises				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Less sticky	Rejected Poor quality Less sticky	Accepted				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
45	Colostomy Bag + Wafer + Clamp	Size 55 / 70, Rate/Set	Muller & Phipps Pakistan Natura Conva Tec	Husnain Pharma Hollister Conform 2	Tanveer Enterprises Alterna Colostomy Bag	Friends Enterprises Bao Health			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Poor material Difficult application	Rejected Poor material Poor clamp			
46	Paste for Colostomy Bag	Rate/No.	Muller & Phipps Pakistan Stomathesive Past Conva Tec	Husnain Pharma Adapt Paste	Tanveer Enterprises Coloplast Paster	Friends Enterprises Bao Health			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected	Sample not provided			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
47	Three way Stopper	with tubing All sizes Rate/No	Friends Enterprises Safety	Ibrahim Enterprises NATIONAL	Husnain Pharma Three way Stoper with Tubing	Tanveer Enterprises Kawa 3 Way Stop Cock	Tahir Pharma Medicare		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Material Stop cock is not working properly	Rejected Poor functionality Poor quality	Rejected Poor functionality Poor quality	Accepted	Rejected Poor functionality Poor quality material		
48	Cord Clamp (Double lock)	Sterilized, Individually Packed Rate/No	Friends Enterprises BELLY KNOT	Ibrahim Enterprises CLASSIC	Al-Hamd Enterprises Cikina	Unisa Pvt. Ltd Unicord	Tahir Pharma Medicare		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Loose lock	Accepted	Rejected Loose lock	Rejected Low quality Loose lock		
49	Bandage Plaster of Paris B.P.C	15cmx2.7Metre, Pack of 12 Rate/No	Friends Enterprises ORTHOPLAST	Hope Enterprises Arsona	Essity Pakistan Ltd Gypsona	Tanveer Enterprises Medi Plast			
	Product Evaluation by TEC / End user.		Rejected Poor quality	Sample not provided	Accepted	Rejected			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected					Poor quality			
50	Cotton Bandage B.P.C	6.5 cm x 6 M, Pack of 12 Rate/No.	Friends Enterprises Cotton Craft Cotton Bandage	Ibrahim Enterprises PRIME BANDAGE	Faisal Pharmaceutical Industries Surgical Bandages	Tanveer Enterprises Medi Band			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Low absorbance	Rejected Poor quality Light weight Low absorbance	Rejected Poor quality Low thread Light weight	Accepted			
51	Cotton Crepe Bandage B.P.C	15cmx4.5 M (Stretched), Pack of 12 Rate/No	Ibrahim Enterprises ELASTOMER	Essity Pakistan Ltd Elastocrepe	Zahid Traders Crepe bandage Faisal Pharma	Tanveer Enterprises Medi Crape	Friends Enterprises Elastocraft		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted	Rejected Poor quality	Rejected Poor quality	Sample not provided		

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
52	Cotton Crepe Bandage B.P.C	7.5 cm x 4.5 M (Stretched), Pack of 12 Rate/No	Ibrahim Enterprises ELASTOMER	Essity Pakistan Ltd Elastocrepe	Zahid Traders Crepe bandage Faisal Pharma	Tanveer Enterprises Medi Crape	Friends Enterprises Elastocraft		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted	Rejected Poor quality	Rejected Poor quality	Sample not provided		
53	Crepe Bandage BPC Roll	Stretched 10cmx4.5m Rate/No.	Ibrahim Enterprises ELASTOMER	Essity Pakistan Ltd Elastocrepe	Tanveer Enterprises Medi Crape	Friends Enterprises Elastocraft			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted	Rejected Poor quality	Sample not provided			
54	Absorbent Cotton Wool B.P.C	500 gm Rate/No	Ibrahim Enterprises LEADER	Tanveer Enterprises Medi Cot					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Poor quality Fiber shredding					
55	Surgical Gauze B.P.C.	100cm x 30M, (Roll) Rate/Meter	Friends Enterprises PRO-MED ROLL	Ibrahim Enterprises KOHINOOR ABSORBENT GAUZE ROLL	Faisal Pharmaceutical Industries Steriband	Tanveer Enterprises Soft Gauze			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality	Rejected Low quality	Accepted	Rejected Low quality			
56	Sterilized Gauze Pads	10cm x 10cm, 8 ply, Rate/10 pieces	Friends Enterprises PROMED	Ibrahim Enterprises MEDI GAUZE	Essity Pakistan Ltd Propax	Zahid Traders Steriband Faisal Pharma	Tanveer Enterprises MediPad		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality	Rejected Low quality	Accepted	Accepted	Rejected Low quality		

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
57	Caston Cotton	6" Rate/No.	Friends Enterprises ORTHOBAN	Ibrahim Enterprises CLASSIC	Essity Pakistan Ltd Soffban	Zahid Traders Sterimax Faisal Pharma	Tanveer Enterprises Ortho Cot		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
58	Polyurethane Dressing	10cm x 12/14cm Rate/No	Saru International Surgical Incise Drape	Essity Pakistan Ltd Opsite Smith & Nephew China	Tanveer Enterprises Surgi-Site				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Quality not upto the mark	Accepted	Rejected Quality not upto the mark				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
59	Chlorhexidine Gauze dressing	15 x 20 cm, Sterilized, Pouch Pack	Friends Enterprises PARA-TULLE	Essity Pakistan Ltd Bactigras Smith & Nephew China	Tanveer Enterprises Septigrass				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Quality not upto the mark	Accepted	Rejected Quality not upto the mark				
60	Wide Area adhesive Roll	10cm x 10 Meter Hypoallergenic Rate/pack	Friends Enterprises PROTECT FIX	Ibrahim Enterprises Recoo	Hope Enterprises FIXOPORE	Saru International Fixing Roll	Essity Pakistan Ltd Fixomull BSN Germany	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Hospital & Homecare
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor sticking	Rejected Poor sticking	Rejected Poor sticking	Rejected Poor sticking	Accepted	Accepted	Rejected Poor sticking
			Tanveer Enterprises						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Oxymax Rejected Poor sticking						
61	Surgical Paper tape (Hypo-allergic)	1" x 5 yards (4.5M) Box of 12 Rate/Dozen	Friends Enterprises HYNAUT	Hope Enterprises ARSOPORE	Al-Hamd Enterprises Kohlcan	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma Hospital & Homecare	Tanveer Enterprises Surgi-Tape	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Poor sticking Low quality paper	Rejected Poor sticking Low quality paper	Accepted	Rejected Poor sticking Low quality materilal	Rejected Poor sticking Low quality paper	
62	I.V. Cannula No. 18	with injection port with wing & stopper Rate/No	Ibrahim Enterprises Green	Al-Hamd Enterprises Inpharven	G.B. Trading Searle	Husnain Pharma Vasofix	Tahir Pharma B.Cat2		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality material Shorter in length	Accepted	Rejected Low quality material Shorter in length	Accepted	Accepted		

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
63	I.V. Cannula No. 20	with injection port with wing & stopper Rate/No	Friends Enterprises VEINOFIX	Ibrahim Enterprises Green	Al-Hamd Enterprises Inpharven	Superior Healthcare International Fero-Cath	Zainabain Enterprises Faisalabad Golden	Husnain Pharma Vasofix	Tanveer Enterprises I Cath
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality material Shorter in length	Rejected Low quality material Shorter in length	Accepted	Accepted	Rejected Low quality material Shorter in length	Accepted	Rejected Low quality material Shorter in length
			Tahir Pharma Medicare						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality material Shorter in length						
64	I.V. Cannula No. 22	with injection port with wing + stopper Rate/No	Friends Enterprises VEINOFIX	Ibrahim Enterprises Green	Al-Hamd Enterprises	Superior Healthcare International Fero-Cath	Zainabain Enterprises Faisalabad Golden	G.B. Trading Searle	Husnain Pharma Vasofix

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
					Inpharven				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality material Shorter in length	Rejected Low quality material Shorter in length	Accepted	Accepted	Rejected Low quality material Shorter in length	Rejected Low quality material Shorter in length	Accepted
			Tanveer Enterprises	Tahir Pharma Medicare					
			I Cath						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Low quality material Shorter in length	Rejected Low quality material Shorter in length					
65	I.V. Cannula No. 24	with stopper and without injection port Rate/No	Ibrahim Enterprises JRZ	Al-Hamd Enterprises Farcocathe	Superior Healthcare International Fero-Cath	G.B. Trading Searle	Husnain Pharma Introcan	Tahir Pharma B.Cat2	Zainabain Enterprises Faisalabad Golden
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Rejected Poor quality	Accepted	Rejected Poor quality	Accepted	Rejected Poor quality	Rejected Poor quality
66	Blood Transfusion Bag	450 ml (Single) Rate/No	Ibrahim Enterprises	Pioneer Pharma M.I	Zainabain Enterprises Faisalabad	G.B. Trading Wego	Husnain Pharma JMS	Tanveer Enterprises DEMOTEK	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			CLASSIC PLUS		STAR BLOOD BAG				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor sticking of sticker Poor quality bag	Rejected Poor sticking of sticker Poor quality material	Rejected Poor sticking of sticker Poor quality bag	Accepted	Sample not provided	Accepted	
67	Blood Transfusion I.V. Set	Rate/No.	Ibrahim Enterprises ST CLASSIC	Pioneer Pharma M.I	Zainabain Enterprises Faisalabad STAR B.T. SET	G.B. Trading Wego	Husnain Pharma JMS	Tanveer Enterprises GB Blood Tranfusion Set	Tahir Pharma Unison
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Filter quality not good Kinks easily	Rejected Filter quality not good Kinks easily	Rejected Filter quality not good Kinks easily	Accepted	Sample not provided	Accepted	Accepted
68	Black Silk No.1	with 30-40 mm 1/2 CRB needle. Minimum Length 75 cm. Foil packing Rate/No.	Friends Enterprises PERFECT FINE SILK BRAIDED	Ibrahim Enterprises B.T. SILK SUTURE	Sind Medical Store Deme Silk USA	Merlin International Surgi Line Vikor Health Care Pakistan	Muller & Phipps Pakistan Sutures Non-Absorb Silk Johnson & Johnson	Zainabain Enterprises Faisalabad Pasilk	Popular International Pvt. Ltd SOF SILK™ SIZE 1 35mm, ½ Circle , Taper Point, 75cm Covidien Medtronic USA
	Product Evaluation by TEC / End user.		Rejected Not according to specification	Rejected Poor quality neelde	Accepted	Accepted Conditionally provided the	Accepted With conditional to	Not according to specification	Sample not provided

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Provided poly propylene instead of black silk			proper size of needle i.e. 40mm needle	provide 40mm needle	Provided cutting needle instead of round body	
			Husnain Pharma Glysilk	Tanveer Enterprises Trusilk	Tahir Pharma Wego				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted With conditional to provide 40mm needle	Accepted	Rejected Poor quality thread & material Small needle				
69	Black Silk No.1	Cutting Needle, Pack of 12 Rate/No.	Friends Enterprises RT MED	Zainabain Enterprises Faisalabad Pasilk	Popular International Pvt. Ltd SOFSILK™ SIZE 1 39mm , 75cm , 3/8 Circle , Reverse Cutting Covidien Medtronic USA	Husnain Pharma Glysilk	Tanveer Enterprises Trusilk	Tahir Pharma Wego	

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Not according to specification Provided 3.0 in pack (each foil) while on pack size 1 is mentioned	Accepted With conditional to provide 40mm needle	Sample not provided	Accepted With conditional to provide 40mm needle	Accepted With conditional to provide 40mm needle	Accepted With conditional to provide 40mm needle	
70	Black Silk No. 2	with 40 mm 1/2 CRB needle. Minimum Length 75 cm. Foil packing Rate/ No.	Friends Enterprises RT MED	Ibrahim Enterprises B.T. SILK SUTURE	Sind Medical Store Deme Silk USA	Merlin International Surgi Line Vikor Health Care Pakistan	Muller & Phipps Pakistan Sutures Non-Absorb Silk Johnson & Johnson	Zainabain Enterprises Faisalabad Pasilk	Husnain Pharma Glysilk
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tanveer Enterprises Trusilk	Tahir Pharma Wego					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
71	Black Silk No. 2/0	with 25-30mm 1/2 CRB needle. Minimum Length 75 cm. Foil packing Rate/ No.	Friends Enterprises PERFECT FINE SILK BRAIDED	Ibrahim Enterprises B.T. SILK SUTURE	Muller & Phipps Pakistan Sutures Non-Absorb Silk Johnson & Johnson	Zainabain Enterprises Faisalabad Pasilk	Popular International Pvt. Ltd SOFSILK™ SIZE 2-0 25mm, Taper Point, 1/2 Circle, 75cm Covidien Medtronic USA	Husnain Pharma Glysilk	Tanveer Enterprises Trusilk
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma						
			Wego						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
72	Black Silk No. 2/0	with 55-65mm cutting needle. Minimum Length 45 cm. Foil packing Rate/ No.	Ibrahim Enterprises B.T. SILK SUTURE	Zainabain Enterprises Faisalabad Pasilk	Popular International Pvt. Ltd SOFSILK , SIZE 2-0 60mm, 75cm, Straight, Conventional Cutting. Covidien Medtronic USA	Husnain Pharma Glysilk	Tahir Pharma Wego		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
73	Black Silk No. 4/0	with 13-17 mm 1/2 CRB needle. Minimu Length 60 cm. Foil packing Rate/ No.	Ibrahim Enterprises B.T. SILK SUTURE	Zainabain Enterprises Faisalabad Pasilk	Husnain Pharma Glysilk	Tahir Pharma Wego			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Not according to specification Provided cutting instead of ½ CRB	Sample not provided	Accepted			
74	Ch. Catguat No.1	with 40 mm ½ CRB needle. Minimum Length 75 cm. Foil packing, Pack of 12 Rate/No.	Friends Enterprises PERFECT FINE CHROMIC CATGUT	Ibrahim Enterprises B.T. CHROMIC CATGUT	Sind Medical Store Deme Gut USA	Merlin International Surgi Line Vikor Health Care Pakistan	Husnain Pharma Chromic Catgut	Tahir Pharma Wego	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
75	Ch. Catguat No.2	with 40 mm ½ CRB needle. Minimum Length 75 cm. Foil packing, Pack of 12 Rate/No.	Friends Enterprises PERFECT FINE CHROMIC CATGUT	Ibrahim Enterprises B.T. CHROMIC CATGUT	Sind Medical Store Deme Gut USA	Merlin International Surgi Line Vikor Health Care Pakistan	Husnain Pharma Chromic Catgut	Tahir Pharma Wego	
	Product Evaluation by TEC / End user.		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms							
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected									
76	Ch. Catguat No.2/0	with 30 mm ½ CRB needle. Minimum Length 75 cm. Foil packing, Pack of 12 Rate/No.	Ibrahim Enterprises B.T. CHROMIC CATGUT	Sind Medical Store Deme Gut USA	Merlin International Surgi Line Vikor Health Care Pakistan	Husnain Pharma Chromic Catgut	Tahir Pharma Wego			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending							
77	Poly glycolic Acid/Polyglactin Braided, Synthetic, Absorbable suture No. 1	with 30mm 1/2 circle Taper cut needle. Minimum Length 75cm. Foil Packing, Pack of 36 or less Rate/No.	Friends Enterprises RT MED	Ibrahim Enterprises B.T. PGA SUTURE	Sind Medical Store Deme Sorb USA	Muller & Phipps Pakistan Sutures Absorbable Vicryl Plus Johnson & Johnson	Zainabain Enterprises Faisalabad STAR	Popular International Pvt. Ltd POLYSORB, SIZE 1 30mm, 1/2 Circle, 75cm Covidien Medtronic USA	Husnain Pharma NOVOSYN VIOLET	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Accepted	Accepted	Accepted	Sample not provided	Accepted With condition to provide extra suture due to 5cm shorter length	

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma						
			Wego						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
78	Poly glycolic Acid/Polyglactin Braided, Synthetic, Absorbable suture No. 2	with 45-50mm, 1/2 CRB Needle, minimum length 75cm Length Rate/No.	Friends Enterprises RT MED	Ibrahim Enterprises B.T. PGA SUTURE	Muller & Phipps Pakistan Sutures Absorbable Vicryl Plus Johnson & Johnson	Zainabain Enterprises Faisalabad STAR	Popular International Pvt. Ltd POLYSORB, SIZE 2 48mm, 1/2 Circle, Taper, 75cm Covidien Medtronic USA	Husnain Pharma Glytin	Tanveer Enterprises Trucryl
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Sample not provided	Not according to specification Provided 40mm needle instead of 50mm	Sample not provided	Not according to specification Provided 40mm needle instead of 50mm	Not according to specification Provided 40mm needle instead of 50mm

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma						
			Wego						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
79	Poly glycolic Acid/Polyglactin Braided, Synthetic, Absorbable suture No. 2/0	with 25-40 mm, 1/2 CRB/Taper Point needle. Minimum Length 75cm. Foil Packing, Pack of 36 or less Rate/No.	Friends Enterprises PERFECT FINE	Ibrahim Enterprises B.T. PGA SUTURE	Sind Medical Store Deme Sorb USA	Muller & Phipps Pakistan Sutures Absorbable Vicryl Plus Johnson & Johnson	Zainabain Enterprises Faisalabad STAR	Popular International Pvt. Ltd POLYGLACTIN, 2-0, 26MM, 1/2 CIRCLE, TAPER POINT, 70CM, VIOLET Covidien Medtronic USA	Husnain Pharma NOVOSYN VIOLET
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Thread is of non-Uniform thickness poor quality	Accepted	Accepted With the condition to provide extra thread 5cm shorter length	Accepted	Sample not provided	Accepted With the condition to provide extra thread 5cm shorter length

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tanveer Enterprises Trucryl	Tahir Pharma Wego					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
80	Poly glycolic Acid/Polyglactin Braided, Synthetic, Absorbable suture No. 3/0	with 20-30 mm, 1/2 CRB/Taper Point needle. Minimum Length 75cm. Foil Packing, Pack of 36 or less Rate/No.	Friends Enterprises PERFECT FINE	Ibrahim Enterprises B.T. PGA SUTURE	Sind Medical Store Deme Sorb USA	Muller & Phipps Pakistan Sutures Absorbable Vicryl Plus Johnson & Johnson	Zainabain Enterprises Faisalabad STAR	Popular International Pvt. Ltd POLYGLACTIN, 3-0, 26MM, 1/2 CIRCLE, TAPER POINT, 70CM, VIOLET Covidien Medtronic USA	Husnain Pharma Glytin
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Thread is of non-uniform thickness Poor quality	Accepted	Sample not provided	Accepted	Sample not provided	Accepted

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tanveer Enterprises Trucryl	Tahir Pharma Wego					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
81	Polyglactin / Poly glycolic Acid Braided, Synthetic, Absorbable suture No. 4/0	with 16-20 mm, 1/2 CRB/Taper Point needle. Minimum Length 75cm. Foil packing, Pack of 36 or less Rate/No.	Friends Enterprises RT MED	Ibrahim Enterprises B.T. PGA SUTURE	Sind Medical Store Deme Sorb USA	Muller & Phipps Pakistan Sutures Absorbable Vicryl Johnson & Johnson	Zainabain Enterprises Faisalabad STAR	Popular International Pvt. Ltd POLYSORB, SIZE 4-0, 17mm, 1/2 Circle, Taper Point, 75cm. Covidien Medtronic USA	Husnain Pharma Glytin
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Sample not provided	Accepted	Sample not provided	Accepted

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma						
			Wego						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
82	Polyamide 10/0 (for eye)	with 6.5-7mm, 3/8 circle micro point spatula double needle, minimum length 30cm. Foil packing, pack of 12 Rate/No.	Friends Enterprises RT MED	Muller & Phipps Pakistan Sutures None Absorb Ethilon Johnson & Johnson	Tahir Pharma Wego				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Needle is too fragile and thread breaks	Accepted	Needle is too thin, fragile and suture breaks while suturing				
83	Poly Propylene No. 1	with 30-40mm, ½ CRB /taper cut needle, minimum length 75cm. Foil packing, pack of 36 or less Rate/No.	Friends Enterprises PERFECT FINE	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Sind Medical Store Deme Lene USA	Merlin International Surgi Lene Vikor Health Care Pakistan	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson	Zainabain Enterprises Faisalabad PALENE	Popular International Pvt. Ltd SURGIPRO II™ SIZE 1 40mm, ½ Circle, Taper Point , 75cm, Blue Covidien Medtronic USA
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Non-uniform thickness & thin thread	Rejected Thikness of thread is non-uniform	Accepted	Accepted	Accepted	Rejected Thikness of thread is non-uniform Low quality material	Sample not provided

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Husnain Pharma OPTILENE	Tanveer Enterprises Trulene	Tahir Pharma Wego				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Thread kinks poor quality material	Accepted				
84	Poly Propylene No. 1	with 40-50mm, cutting needle, minimum length 75cm. Foil packing, Pack of 36 or less Rate/No.	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Zainabain Enterprises Faisalabad PALENE	Husnain Pharma OPTILENE	Tahir Pharma Wego			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending Required specification with 60mm needle, curved cutting needle, minimum length 75cm						
85	Poly Propylene No. 2/0	with 25-40mm, ½ CRB needle, minimum length 75cm. Foil packing, Pack of 36 or less Rate/No.	Friends Enterprises PERFECT FINE POLYPROPYLENE	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Sind Medical Store Deme Lene USA	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson	Zainabain Enterprises Faisalabad PALENE	Popular International Pvt. Ltd SURGIPRO II™ SIZE 2-0 25mm, 1/2 Circle, Taper Point, 75cm, Double Ended Needle Covidien Medtronic USA	Husnain Pharma OPTILENE

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Suture not having uniform thickness	Rejected Poor quality Suture	Accepted	Accepted	Rejected Poor quality Suture	Sample not provided	Accepted
			Tanveer Enterprises Trulene	Tahir Pharma Wego					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
86	Poly Propylene No. 2/0	with 25-40mm, cutting needle, minimum length 75cm. Foil packing, Pack of 36 or less Rate/No.	Friends Enterprises RT MED	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Sind Medical Store Deme Lene USA	Merlin International Surgi Lene Vikor Health Care Pakistan	Zainabain Enterprises Faisalabad PALENE	Husnain Pharma Glylene	Tahir Pharma Wego
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Material of thread is non-uniform	Accepted	Accepted	Rejected Material of thread is non-uniform	Accepted	Accepted

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
87	Polypropylene No. 3/0	with 16-25mm, 1/2 CRB/Taper Point needle, minimum length 75cm. Foil packing Rate/No.	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Sind Medical Store Deme Lene USA	Merlin International Surgi Lene Vikor Health Care Pakistan	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson	Zainabain Enterprises Faisalabad PALENE	Popular International Pvt. Ltd SURGIPRO II™ SIZE 3-0 17mm, 75cm, 1/2 Circle, Taper Point, Double Needled Covidien Medtronic USA	Husnain Pharma Glylene
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Thin thread Non-uniform	Accepted	Accepted	Accepted	Rejected Poor quality thread	Sample not provided	Rejected Needle quality is poor
			Tahir Pharma Wego						
	Product Evaluation by TEC / End user.		Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
88	Polypropylene No. 4/0	with 15-25mm, ½ CRB needle, minimum length 75cm. Foil packing, Pack of 36 or less Rate/No.	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Sind Medical Store Deme Lene USA	Merlin International Surgi Lene Vikor Health Care Pakistan	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson	Zainabain Enterprises Faisalabad PALENE	Popular International Pvt. Ltd SURGIPRO II™ SIZE 4-0 25mm, 75cm, 1/2 Circle, Taper Point, Double Needled Covidien Medtronic USA	Husnain Pharma OPTILENE
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Accepted	Rejected Thread quality not upto the mark	Sample not provided	Sample not provided
			Tanveer Enterprises Trulene	Tahir Pharma Wego					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
89	Poly Propylene 6/0	Double needle Round body Rate/No.	Ibrahim Enterprises B.T POLYPROPYLENE SUTURE	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson	Husnain Pharma OPTILENE	Tahir Pharma Wego			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted	Accepted	Accepted			
90	Bone Wax	Rate/Dozen	Popular International Pvt. Ltd BONE WAX 2.5G, Cylinder Shape Covidien						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
91	Disposable syringe (Blister Pack)	50 cc, with side needle Rate/No	Ibrahim Enterprises CLASSIC FINE	Hope Enterprises CURE	Superior Healthcare International Star	Husnain Pharma UCI	J.J. Brothers Enterprises Golden	Tahir Pharma Shifa	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected No leur lock	Sample not provided	Rejected No leur lock	Sample not provided	Accepted	Sample not provided	
92	Disposable syringe (Blister Pack)	50/60 cc, with Central nozzle or catheter tip Rate/No	Ibrahim Enterprises CLASSIC FINE	Superior Healthcare International Star	Husnain Pharma UCI	J.J. Brothers Enterprises Golden	Tanveer Enterprises Hospital & Homecare Disposable Syringe	Tahir Pharma Shifa	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Poor material Difficult sliding material of plunger	Accepted	Poor material Difficult sliding material of plunger	Accepted	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
93	Disposable syringe with needle (Blister Pack)	20cc, With Central Nozzle with Leur lock Rate/No	Ibrahim Enterprises CLASSIC FINE	Unisa Pvt. Ltd Uniject	Superior Healthcare International Star	Zainabain Enterprises Faisalabad BM DISPOSABLE	J.J. Brothers Enterprises Golden	Tanveer Enterprises Hospital & Homecare Disposable Syringe	Tahir Pharma Shifa
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Syringe plunger slides with difficulty	Rejected Hard plunger	Accepted	Rejected Poor quality material	Rejected Poor quality material	Accepted
94	Disposable Syringe 20cc	Side Nozle with Needle Rate/No.	Ibrahim Enterprises CLASSIC FINE	Sind Medical Store Yushou China	Husnain Pharma UCI	Tahir Pharma Shifa			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected	Accepted			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
95	Disposable syringe with needle (Blister Pack)	10 cc , with Leur Lock . 21G needle. Rate/No	Ibrahim Enterprises SMD	Sind Medical Store Yushou China	Unisa Pvt. Ltd Uniject	Superior Healthcare International Star	Zainabain Enterprises Faisalabad BM DISPOSABLE	Husnain Pharma UCI	Tanveer Enterprises Apple Disposable Syringe
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Rejected Plunger slide not smooth	Accepted	Accepted	Rejected Poor quality material plunger poor	Accepted
			Tahir Pharma Bio Safe						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality material Plunger slide not smooth						
96	Disposable syringe with needle (auto disposable syringe) (Blister Pack)	5 cc , with leur lock and 23G needle Rate/No	Friends Enterprises LD TRU LOCK	Ibrahim Enterprises SMD	Hospital Services & Sales Haiou	Unisa Pvt. Ltd Unilock	Zainabain Enterprises Faisalabad BM DISPOSABLE	Husnain Pharma YI Xin AD Syringe	Tanveer Enterprises Apple K1 5cc Syringe

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Sample not provided	Rejected Poor quality material plunger quality not smooth	Rejected Poor quality material plunger slide not smooth	Accepted	Rejected Non-functional lock of auto disable syringe
			Tahir Pharma Bio Safe						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality material						
97	Insulin syringe (Blister Pack)	1 cc with needle 30 G, length of needle 5/16 inch Rate/No	Ibrahim Enterprises CLASSIC FINE	Sind Medical Store Yushou China	Superior Healthcare International Star	Tahir Pharma Shifa			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Material poor Hard plunger	Accepted	Accepted			
98	Butter fly needle	24 Size Rate/No, UK, USA Japan, EU origin only. Rate/No	Husnain Pharma JMS	Tanveer Enterprises Nipro Scalp Vein Set					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
99	I.V Infusion set	Rate/No.	Friends Enterprises INFUCAN	Ibrahim Enterprises NATIONAL	Health Linkers Orbit	Unisa Pvt. Ltd Uniset	Superior Healthcare International Star	Zainabain Enterprises Faisalabad BM DISPOSABLE	G.B. Trading Wego
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted	Rejected Poor quality	Rejected Poor quality	Accepted	Rejected Poor quality	Accepted
			J.J. Brothers Enterprises Golden	Tahir Pharma Shifa					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided					
100	Disposable Infusion Set for giving fluids (Burette Chamber)	100 ml Rate/No	Ibrahim Enterprises CLASSIC FINE	Hope Enterprises CURE	Pioneer Pharma Relation	Al-Hamd Enterprises	Hope Pharma	Superior Healthcare International	Zainabain Enterprises Faisalabad

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
						Cikina	BM Bettermost	Star	WELLMED
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Kinks easily	Sample not provided	Rejected Kinks easily	Rejected Kinks easily	Rejected Kinks easily	Rejected Kinks easily	Accepted
			G.B. Trading Wego	Husnain Pharma Hospital & Homecare	J.J. Brothers Enterprises Golden	Tanveer Enterprises Care Burette	Tahir Pharma Shifa		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected	Rejected	Accepted	Accepted		
101	I.V Chamber Burette Set	150 ml Rate/No.	Hope Enterprises CURE	G.B. Trading Wego					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
102	Examination Gloves, Rubber Latex, Powdered.	All Sizes, Box of 100 (By Weight). Rate / Box	Friends Enterprises Safety	Ibrahim Enterprises HIGH MAX	Eastern Medical Care ASAP Malaysia	Shams Scientific Traders Maxlife	Al-Hamd Enterprises Optishield	Superior Healthcare International Senior	Zainabain Enterprises Faisalabad WELLMED
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Rejected Difficult to wear Tears easily	Rejected Tears easily Difficult to wear	Accepted	Accepted
			G.B. Trading Raysen	Tanveer Enterprises Handex					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected	Accepted					
103	Polythene Gloves	Pack of 100	Ibrahim Enterprises MOLLY DALE	Zainabain Enterprises Faisalabad WELLMED	J.J. Brothers Enterprises				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Thin material	Rejected Poor quality Thin material	Accepted				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
104	Surgical Gloves No. 6.5 Min. Length (290mm ± 10mm) (Min. Width Size 6.5, 83mm ± 6mm. Weight 18.5gm and above (without envelope) will be preferred	Powdered Rubber latex surgical gloves sterilized. Rate/Pair	Friends Enterprises SUPERMAX	Ibrahim Enterprises HIGH MAX	Al-Hamd Enterprises Surgitex	Noor International Intouch powdered	Superior Healthcare International Safety	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Raysen
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor rubber quality	Rejected Poor rubber quality & packing	Rejected	Accepted	Rejected	Rejected	Sample not provided
			Husnain Pharma Ansell	Tahir Pharma Maxitex					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Rejected Rubber quality very poor Discoloured rubber					
105	Surgical Gloves No. 7. Min. Length (290mm ± 10mm) (Min. Width size. 7, 89mm ± 6mm. Weight 19.5gm and above (without envelope) will be preferred	Powdered Rubber latex surgical gloves sterilized. Rate/Pair	Friends Enterprises SUPERMAX	Ibrahim Enterprises HIGH MAX	Al-Hamd Enterprises Surgitex	Noor International Intouch powdered	Superior Healthcare International Safety	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Raysen
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor rubber quality	Rejected Poor rubber quality & packing	Rejected Poor rubber quality	Accepted	Rejected Poor rubber quality	Rejected Poor rubber quality	Sample not provided

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma Medicare						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided						
106	Surgical Gloves No. 7.5 Min. Length (290mm ± 10mm) (Min. Width Size 7.5 90mm ± 6mm. Weight 21gm and above (without envelope) will be preferred	Powdered Rubber latex surgical gloves sterilized. Rate/Pair	Friends Enterprises SUPERMAX	Ibrahim Enterprises HIGH MAX	Al-Hamd Enterprises Surgitex	Noor International Intouch powdered	Superior Healthcare International Safety	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Raysen
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor rubber quality	Rejected Poor rubber quality & packing	Rejected Poor rubber quality	Accepted	Rejected Poor rubber quality	Rejected Poor rubber quality	Rejected Poor rubber quality & packing

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Tahir Pharma Medicare						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor rubber quality & packing						
107	Surgical Gloves No. 8 Min. Length (290mm ± 10mm) (Min. Width Size 8, 102mm) ± 6mm. Weight 22gm and above (without envelope) will be preferred	Powdered Rubber latex surgical gloves sterilized. Rate/Pair	Friends Enterprises SUPERMAX	Ibrahim Enterprises HIGH MAX	Al-Hamd Enterprises Surgitex	Noor International Intouch powdered	Superior Healthcare International Safety	Zainabain Enterprises Faisalabad WELLMED	G.B. Trading Raysen
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor rubber quality	Rejected Poor rubber quality & packing	Rejected Poor rubber quality	Accepted	Rejected Poor rubber quality	Rejected Poor rubber quality	Sample not provided
			Tahir Pharma Maxitex						
	Product Evaluation by TEC / End user.		Rejected						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Poor rubber quality & discoloured						
108	Folly's Catheter (Two Way)	08 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead	Tahir Pharma (Chroma Brand)	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Bulge out balloon	Not according to specification	No internal covering No stylet	Accepted	Rejected Bulge out balloon	Sample not provided	
109	Folly's Catheter (Two Way)	10 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	No internal covering	Accepted	Accepted		

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
110	Folly's Catheter (Two Way)	12 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Not according to specification 10fr	No internal covering	Rejected Bulge out balloon	Rejected Bulge out balloon		
111	Folly's Catheter (Two Way)	14 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted	No internal covering	Accepted	Accepted		
112	Folly's Catheter (Two Way)	16 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises RECOO	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead	Tahir Pharma (Chroma Brand)	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	No internal covering Not present	Accepted	Accepted	Sample not provided	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
113	Folly's Catheter (Two Way)	18 No. (Rubber Latex / Silicon Treated) Rate/No.	Friends Enterprises PERFECT	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad Pigeon	Husnain Pharma U- Flow	J.J. Brothers Enterprises Well Lead		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Hard balloon bulge No internal covering	Accepted	Accepted		
114	Folly's Catheter (Three Way)	22 No.	N/Q						
115	Folly's Catheter (Three Way)	24 No. (Rubber Latex / Silicon Treated) Rate/No.	Ibrahim Enterprises KOSAN ANGEL	J.J. Brothers Enterprises Well Lead					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Leackage of irrigation Poor drainage					
116	Spinal Needle /L.P Needle	16 No. Rate/No.	N/Q						
117	Spinal Needle /L.P Needle	23 No.Rate/No.	Husnain Pharma Egemen	Tanveer Enterprises Nipro Spinal needle					
	Product Evaluation by TEC / End user.		Accepted	Accepted					

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
118	Spinal Needle / L.p Needle	25 No.Rate/No.	Friends Enterprises BD	Noor International Uniever	Husnain Pharma Egemen	Tanveer Enterprises Nipro Spinal needle			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Accepted			
119	Spinal Needle /L.P Needle	27 No. Rate/No.	Noor International Uniever	Husnain Pharma Egemen	Tanveer Enterprises Nipro Spinal needle				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
120	Haemodialysis Catheter Kit	12F X 15 cm Double Lumen	Flowtronix Systems Able	Popular International Pvt. Ltd GM-Key	Tahir Pharma Medcomp				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Bad user experience				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
121	Haemodialysis Catheter Kit	12F X 16cm Double Lumen	Flowtronix Systems Able	Popular International Pvt. Ltd GM-Key	Tanveer Enterprises ARROW				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
122	Long Term Permanent Haemodialysis Catheter kit	Dual lumen Radio Opaque Catheter, 15.5f x 23cm (tip to cuff) End caps tunneler introducer needle 18g x 70mm dilator 10-12f x 15cm dialtor 8f x 20cm Guide wire J tip 3mm radias x 70cm marked Air Guide valved introduced, appeal apart sheath 15f with Vassel dilator	Tahir Pharma Medcomp						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
123	AV Fistula Needle	16, 17 guage, Rate / No.	Fresenius Medical Care Thailand	Popular International Pvt. Ltd Vital Healthcare Malaysia					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Bad user experience					
124	Disinfectant Solution for Dialysis Machine	Compatible with Fresenius Machine, Bottle of 5 liter	Eastern Medical Care Farmasol Turkey	Husnain Pharma Lifemed Citro HD	Tahir Pharma Citroclean				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Bad user experience	Rejected Used in machine No good experience				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
125	Tubing Set for Dialyser compatible with fresenius machine	Rate / No	Ibrahim Enterprises BIOTEQ HEMODIALYSIS TUBING SET	Fresenius Medical Care Fresline Vital Healthcare SDH BHD Malaysia	Popular International Pvt. Ltd Vital Healthcare Malaysia				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted	Accepted				
126	Hollow Fiber Dialyzer Compatible with Fresenius Machines	Min. 1.7m2Rate / No	Ibrahim Enterprises Searle Hollow Fiber Demodialyzer	Fresenius Medical Care Germany	Popular International Pvt. Ltd Vital Healthcare Malaysia				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
127	Filter for Dialysis Machine (polysulfonated)	Compatible with Fresenius Machine Rate/No.	Fresenius Medical Care Germany						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
128	Chest Electrodes	Standard Size, Pack of 50, (Disposable) Rate/No.	Ibrahim Enterprises RECOO	Zainabain Enterprises Faisalabad WELLMED	Husnain Pharma My Heart				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor material Poor gel quality	Accepted	Rejected Poor material Poor sticking				
129	Tracheostomy Tube Size # 2.5, 3, 3.5, 4, 4.5, 5	Disposable without cuff Rate/No.	Popular International Pvt. Ltd Shiley Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided						
130	Tracheostomy Tube Size # 5, 5.5, 6, 6.5, 7, 7.5, 8 No.	Disposable with cuff Rate/No.	Popular International Pvt. Ltd Shiely Covidien Medtronic USA	J.J. Brothers Enterprises Well Lead					
	Product Evaluation by TEC / End user.		Sample not provided	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
131	Ultrasound Gel	Pack of 5kg or Litre with applicator	SAA International Skygel Malaysia	J.J. Brothers Enterprises ECO-Super Gel CeraCarta					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted					
132	Under Water Seal	with connector and Bottle with plastic pipe 1 meter Rate/No	Ibrahim Enterprises MEDILINE	J.J. Brothers Enterprises					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Plastic bottle soft and collapsible connector quality not good					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
133	Ligation clips	(For Laparoscopy), Small / Medium / large size , Cartridge of 6 clips Rate/Cartridge (madeup of Titanium), One Clip Applicator with each 400 clips will be provided by the firm FOC.	M Yousaf & Co. V clip Grena UK	Hoorra Pharma Pvt. Ltd Ligaclip Extra Titanium Ethicon, Inc. Johnson & Johnson	Tahir Pharma Maxi Clip	Note Both companies are directed to demonstrate the small medium & large clip applicators before finalization			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
134	Polysling, Best Quality	Large Size Rate/No	Ibrahim Enterprises BLUE BIRD	Health Linkers Active	J.J. Brothers Enterprises Best Quality	Tahir Pharma Dynamic			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality	Accepted	Rejected	Accepted			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
135	Polysling, Best Quality	Small Size Rate/No	Ibrahim Enterprises BLUE BIRD	Health Linkers Active	J.J. Brothers Enterprises Best Quality	Tahir Pharma Dynamic			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected	Accepted	Rejected	Accepted			
136	Lumbosachral Belt	All Sizes Rate/No	Ibrahim Enterprises BLUE BIRD	Health Linkers Active	J.J. Brothers Enterprises Best Quality	Tahir Pharma Dynamic			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected	Accepted			
137	Cervical Collar	All sizes / Hard chin	Ibrahim Enterprises BLUE BIRD	Health Linkers Active	J.J. Brothers Enterprises Best Quality	Tahir Pharma Dynamic			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Not according to specification	Accepted	Rejected Poor quality Less adjustable	Rejected Poor quality			
138	Stinmen Pin	With Stirup 3.5mm Rate/No	Suriquips	J.J. Brothers Enterprises					

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Best Quality					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Rusted sample	Rejected Poor quality					
139	Stinmen Pin	With Stirup 4.5mm Rate/No	Suriquips	J.J. Brothers Enterprises Best Quality					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Rusted sample	Rejected Poor quality					
140	Speed Track / Skin Traction (Complete Kit)	All sizes Rate/kit	Ibrahim Enterprises SKIN TRACTION KIT	Health Linkers Active	Tahir Pharma Dynamic				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Material is very poor	Accepted	Rejected Material is very poor				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
141	Gigly saw	Two twist , length 24" Rate/No	Ibrahim Enterprises Local	Suriquips	G.B. Trading Local	J.J. Brothers Enterprises Best Quality			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected	Sample not provided	Rejected Wire is very thin			
142	Chest Tube	Sterile Blister Pack, Size CH/FR 24, 28,30,32,36 Rate/No	Friends Enterprises Jiejing	Husnain Pharma Chest tube					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
143	Vaccuum Drain bottle 400ml with tubing (All size)	Sterile, 400ml with connecting tube and one piece adapter for single use Rate/No	Ibrahim Enterprises MEDILINE	J.J. Brothers Enterprises	Tanveer Enterprises Redon Bottle				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Poor quality Difficult to handle Vaccum needs to be maintained through machine	Accepted	Poor quality Difficult to manage Vaccum needs to be maintained through machine				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
144	Heat Moisture Exchange Filter (Sterilized)	Rate / No	Ibrahim Enterprises HITECARE	Noor International Rvent	Zainabain Enterprises Faisalabad WELLMED	Popular International Pvt. Ltd DAR Covidien Medtronic USA	J.J. Brothers Enterprises Well Lead	Tanveer Enterprises Emedical Heat Moisture Exchange Filter	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Sample not provided	Accepted	Accepted	
145	Breathing Filter for HIV / HCV / HBV (Sterilized)	Rate / No	Ibrahim Enterprises HITECARE	Noor International Rvent	Popular International Pvt. Ltd DAR Covidien Medtronic USA	J.J. Brothers Enterprises Well Lead	Tanveer Enterprises Emedical Breathing Filter		
	Product Evaluation by TEC / End user.		Accepted	Accepted	Sample not provided	Accepted	Accepted		

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
146	Dj Stents set 4.8 Fr. Straight Tip 0.032" size (Including Guide wire)	Medium Pressuer, Rate/ No. FDA/CE Approved	Friends Enterprises VENOUS	Hope Enterprises DJS 4.7fr-26cm Both End Open with Guide Wire	J.J. Brothers Enterprises Coloplast	Tahir Pharma Plastemed			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Sample not provided	Accepted	Accepted			
147	Dj Stents set 6.0 Fr. (Including Guide wire)	Medium Pressuer, Rate/ No. FDA/CE Approved	Friends Enterprises Copper	Hope Enterprises DJ Stent 6Fr-26cm Both End Open Both End Multiloop	J.J. Brothers Enterprises Coloplast	Tahir Pharma Plastemed			
	Product Evaluation by TEC / End user.		Accepted	Rejected Guide wire size not mentioned	Accepted	Accepted			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
148	Ureteric catheter	5fr Rate/No. FDA/CE Approved	J.J. Brothers Enterprises Well Lead	Tahir Pharma Plastemed					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Non tapered end Stylet not present	Accepted					
149	Guide Wire J-Tip all sizes (For Urology)	(J-Tip) Rate/No. FDA/CE Approved	Friends Enterprises Copper	Tahir Pharma Ningbo					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted					
150	Guide Wire Straight tip all sizes (For Urology)	(Straight tip) Rate/No. FDA/CE Approved	Friends Enterprises Copper	Tahir Pharma Ningbo					
	Product Evaluation by TEC / End user.		Accepted	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
151	TUR Loops	Adult Size	J.J. Brothers Enterprises						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Wolf Accepted						
152	Urethrotomy Knife	Adult Size	N/Q						
153	PCN Tube	8fr Rate /No.	Hope Enterprises PCN Catheter 8fr	Flowtronix Systems Able	J.J. Brothers Enterprises Coloplast	Tahir Pharma Plastemed			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Not according to specificaiton	Accepted	Accepted			
154	PCNL Drap Sheet	Rate/No	Tanveer Enterprises Medpro PCNL Drape Sheet						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
155	Yaunker Suction Set	Rate/No	Friends Enterprises KB	Ibrahim Enterprises HITECARE	Husnain Pharma Yaunker Suction	J.J. Brothers Enterprises Well Lead	Tahir Pharma Medicare		
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Poor quality pipe Kinks easily	Accepted	Accepted		
156	Phaco Knife 15 degree for eye	Rate / No	Kamran Bashir and Brothers Sharp point	Optisurg Oryum knife					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
157	Phaco Knife 3.2/3.25 mm for eye	Rate / No	Kamran Bashir and Brothers Sharp point	Optisurg Oryum knife					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
158	Phaco Knife 2.7 mm for eye	Rate / No	Kamran Bashir and Brothers Sharpoint	Optisurg Oryum knife					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
159	Intra Ocular Lens + Injector	Foldable All sizes, DRAP Registered	Health Linkers Acriva	Optisurg BiFlex					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
160	Mesh Fixation Device for Laproscopy	Coil Design Tackers or equivalent	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Protack Lap Mesh Fixation					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Device, 30 Coil Design Tackers Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
161	Round Stapler	Tilt top anvil / equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Circular Mech Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Disposable Circular Stapler All Sizes Kangdi a covidien company				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
162	Linear cutter and stapler Reloadable Size. 60/55	Directional stapling technology / 60/55 Long Anvil with cutter and safety Lockout or equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Promixate Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd GIA Stapler With DST Series Technology				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
					60mm (One Reload Included) Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
163	Linear cutter and stapler Reloadable Size. 80/75	Directional stapling technology / 80/75 Long Anvil with cutter and safety lockout or equivalent	M Yousaf & Co. Panther China	Hoorah Pharma Pvt. Ltd Promixate Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd GIA Stapler With DST Series Technology 80mm (One Reload Included) Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
164	Linear cutter and stapler Reloadable Size.100	Directional stapling technology / 100mm Long Anvil with cutter and safety lockout or equivalent	M Yousaf & Co. Panther China	Hoorah Pharma Pvt. Ltd Promixate Ethicon, Inc.	Popular International Pvt. Ltd GIA Stapler With DST Series Technology				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Johnson & Johnson	100mm (One Reload Included) Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
165	Reload of Linear cutter/size. 60/55	Reload for Linear cutter with titanium staples/equivalent. 60/55	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Promixate Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd GIA Loading Unit With DST Series Technology, Size 60mm-3.8mm Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
166	Reload of Linear cutter/size. 80/75	Reload for Linear cutter with titanium staples/equivalent. 80/75	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Promixate Ethicon, Inc.	Popular International Pvt. Ltd GIA Loading Unit With DST Series				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Johnson & Johnson	Technology, Size 80mm-3.8mm Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
167	Reload of Linear cutter/ Proximate Linear Cutter No.100	Reload for Linear cutter with titanium staples/equivalent. 100	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Promixate Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd GIA Loading Unit With DST Series Technology, Size 100mm-3.8mm Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
168	Linear cutter and stapler instrument for EndoSurgery All Sizes	Laposcopic Stapler with 12mm Diameter. Ability to cut and staple forming 3rows on each side of A cutline when Reload is Attached / Equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Articulating Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Endo GIA Stapler, All Size Covidien Medtronic USA				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
169	Reload for endo surgery instrument (Triple Staggered) 30mm	30mm, Reload for laproscopic stapler with 12mm diameter. Ability to cut and staple forming 3rows on each side of a cutline with blade/equivalent	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Endo GIA Reload 30MM , Gray, White ,Blue Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
170	Reload for endo surgery instrument (Triple Staggered) 45mm	45mm, Reload for laproscopic stapler with 12mm diameter. Ability to cut and staple forming 3rows on each side/equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Echelon Ethicon, Inc.	Popular International Pvt. Ltd Endo GIA 45MM Reload , White ,Blue & Green,				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Johnson & Johnson	Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
171	Reload for endo surgery instrument 60mm	60mm, Reload (Straight) for laproscopic stapler with 12mm diameter. Ability to cut and staple forming 3rows on each side/equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Echelon Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Signia Cartridge 60mm Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
172	Reload for endo surgery instrument 30mm	30mm, Reload (Roticulator) for laproscopic stapler with 12mm diameter. Ability to cut and staple	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Endo GIA					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		forming 3rows on each side/equivalent		Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
173	Reload for endo surgery instrument 45mm	45mm, Reload (Reticulator) for laproscopic stapler with 12mm diameter. Ability to cut and staple forming 3rows on each side/equivalent	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Endo GIA Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
174	Reload for endo surgery instrument 60mm	60mm, Reload (Reticulator) for laproscopic stapler with 12mm diameter. Ability to cut and staple	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Endo GIA					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		forming 3rows on each side/equivalent		Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
175	Disposable Vessel Sealing Hand Instrument 10mm (Open Surgery)	Disposable Vessel Sealing Hand Instrument 10mm for Open Surgery / Equivalent. Vessel sealing system & Services will be provided free of cost to the hospital	Popular International Pvt. Ltd Ligasure Atlas Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
176	Vessel Sealing 20cm Hand Switching Laparoscopic Instrument/Equivalent	Vessel Sealing Atlas 20cm Hand Switching Laparoscopic Instruemnt/Equivalent. Vessel sealing system & Services will be	Hoorra Pharma Pvt. Ltd	Popular International Pvt. Ltd Ligasure	176,177,178,182,183 with100 hand pieces generator will be FOC and if quality less then 100, ethicon will place generator for usage.				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		provided free of cost to the hospital.	Enseal Ethicon, Inc. Johnson & Johnson	Covidien Medtronic USA	Condition for both ethicon & covidien				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
177	Vessel Sealing Impact / Enseal Instrument /Equivalent	Vessel Sealing Impact/Enseal Instrument / Equivalent. Vessel sealing system & Services will be provided free of cost to the hospital.	Hoora Pharma Pvt. Ltd Enseal Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Ligasure Impact Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
178	Vessel Sealing Curved 5mm Jaw Sealer / Divider / Equivalent	Vessel Sealing Curved 5mm Jaw Sealer / Divider / Equivalent. Vessel sealing system & Services will be	Hoora Pharma Pvt. Ltd	Popular International Pvt. Ltd					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		provided free of cost to the hospital.	Enseal Ethicon, Inc. Johnson & Johnson	Ligasure Maryland 5mm Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
179	Circular Stapler	2 staggered rows / equivalent	M Yousaf & Co. Panther China	Hoorra Pharma Pvt. Ltd Circular Mech Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Disposable Circular Stapler Kangdi a Medtronic company				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				
180	Curved cutter stapler	40mm equivalent with one reload inside	Hoorra Pharma Pvt. Ltd	Popular International Pvt. Ltd					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Contour Ethicon, Inc. Johnson & Johnson	KFQ Stapler Kangdi a Medtronic company					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
181	Reload for curved cutter stapler	40mm /equivalent	Hoora Pharma Pvt. Ltd Curved Cutter Reload Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd KFQH-40 Stapler Cartridge Kangdi a Medtronic company					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
182	Disposable vessel sealing Hand Instrument / Equivalent (open surgery) different types blunt tip/equivalent curved jaw/equivalent, small jaw/equivalent Large jaw corved tip/equivalent	Disposable Vessel sealing hand instrument for open surgery/ equivalent. Vessel sealing system & Services will be provided free of cost to the hospital	Hoora Pharma Pvt. Ltd Enseal Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Ligasure Hand Instrument Open Sealer Divider Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
183	Vascular surgery stapler	45mm white / equivalent	M Yousaf & Co. Panther China	Hoora Pharma Pvt. Ltd Echelon Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Signia Cartridge 45MM Tan with Tri-Staple Technology Covidien Medtronic USA				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
184	Vascular surgery stapler	30mm tan / equivalent	M Yousaf & Co. Panther China	Popular International Pvt. Ltd ENDO GIA RELOAD 30MM Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
185	Vessel Sealing Curved 5*37mm jaw sealer / divider / Equivalent	Vessel sealing system & services will be provided free of cost to the hospital	Hoorra Pharma Pvt. Ltd Enseal Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd LIGASURE MARYLAND Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
186	Articulating stappler	60mm multi use loading unit / equivalent	M Yousaf & Co. Panther China	Popular International Pvt. Ltd Signia Loading Unit 60mm Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
187	Articulating stappler 45mm multi use loading unit	40mm multi use loading unit / equivalent	Popular International Pvt. Ltd Signia Loading Unit 45mm Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
188	Vessel Sealing hand Instrument for Thyroectomy		Hoora Pharma Pvt. Ltd Focus Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Ligasure Small Curved Jaw Open Sealer Divider Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
189	Strips for Glucometer	Pack of 50, Rate/Strip, Supplier will be bound to provide one hundred Glucometer with supply of one lac strips, free of cost	Mian Scientific Corporation Pvt. Ltd True Answer	G.B. Trading Sejoy	Popular International Pvt. Ltd Oncall Plus Acon				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
190	Blood Lancet Pricker	Rate/No.	Ibrahim Enterprises RECO	Mian Scientific Corporation Pvt. Ltd China	Popular International Pvt. Ltd Oncall Lancet Acon	Tahir Pharma Medicare			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Accepted	Accepted			
191	Pressure tube	Rate/No	Friends Enterprises Extension Set	J.J. Brothers Enterprises SCW MEDICATH	Tahir Pharma Medicare				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Hard tube kinks easily				
192	Forgatory (Embolectomy) Catheter	All sizes Rate/No.	N/Q						
193	Ligation band	Pack of 100	N/Q						
194	Adult Rem polyhesive II patient Return electrode	Patient Return Electrode must be used with present electric cauteries in hospital.	Popular International Pvt. Ltd Adult Rem Polyhesive II Patient Return Electrode Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
195	Patient Return electrode/ Clamp reusable	Patient Return Electrode must be used with present electric cauteries in hospital.	Popular International Pvt. Ltd Clamp Cord Resuable Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
196	Dignity Sheet/Under Pad	Pack 10 sheets, Rate per pack	Ibrahim Enterprises MEDI PLUS	Zainabain Enterprises Faisalabad WELLMED	Tanveer Enterprises Uni Care				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor & thin material	Accepted	Accepted				
197	Absorbable Gelatin Spong with Haemostatic Effect	Standard Size, Rate/No.	Friends Enterprises Cutanplast	Zainabain Enterprises Faisalabad WELLMED	J.J. Brothers Enterprises Monarch				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Hard material Difficult to handle				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
198	Oxidized regenerated cellulose	All Sizes, Rate/No.	Muller & Phipps Pakistan Surgicel Johnson & Johnson	Popular International Pvt. Ltd Traumastm Baxter Healthcare USA	Husnain Pharma Oxicell				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Sample not provided				
199	Skin Stapler	35 pins, Rate/No.	Friends Enterprises Microcure	M Yousaf & Co. Citec Pride China	Zainabain Enterprises Faisalabad WELLMED	Hoor Pharma Pvt. Ltd Proximate Ethicon, Inc. Johnson & Johnson	Popular International Pvt. Ltd Disposable Skin Stapler (Kangdi A Medtronic Company) (Kangdi A Medtronic Company)	Tanveer Enterprises Celtics	
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Handle not smooth Poor quality	Accepted	Accepted	Accepted	Sample not provided	Rejected Clip not approximate effectely Thin pins material	

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
200	Tracheostomy Mask, with tubing complete set	Rate/No.	Tanveer Enterprises Emedical Traechostomy Mask						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
201	True Cut Biopsy Needle	14G & 16G Rate/No. (Semiautomatic)	Friends Enterprises VENOUS	J.J. Brothers Enterprises Monarch	Tahir Pharma True Cut Biopsy Needle				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Difficult to handle	Rejected Difficult to handle				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
202	Mesh Polypropylene 10x15cm	Light Weight Rate/No.	Medi Lution Health Care Beta Tech Turkey	Popular International Pvt. Ltd Parietene Light Weight Mesh (Poly Propylene) , Size 10x15cm, Covidien Medtronic USA					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided	Accepted					
203	Mesh Polypropylene 15x15cm	Light Weight Rate/No.	Hope Enterprises Poly Propylene Sietka Mesh Light weight	Medi Lution Health Care Beta Tech Turkey	Popular International Pvt. Ltd Parietene Light Weight Mesh , (Poly Propylene) , Size 15x15cm, Covidien Medtronic USA	J.J. Brothers Enterprises Duzey Medical			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Sample not provided	Accepted	Rejected Hard mesh			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
204	Polypropylene Mesh	Small (6x11cm/8x13cm) Rate/No.	Hope Enterprises Poly Propylene Sietka Mesh Non absorbable Mesh Undyed Size	Medi Lution Health Care Beta Tech Turkey	Popular International Pvt. Ltd Parieten Macroporus Mesh , (Poly Propylene) , Size 11x06cm , Covidien Medtronic USA	J.J. Brothers Enterprises Duzey Medical			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Poor quality Material Difficult to handle	Accepted	Rejected Poor quality Material Difficult to handle			
205	Polypropylene Mesh	Large (15x15cm) Rate/No.	Hope Enterprises Poly Propylene Sietka Mesh Non absorbable Mesh Undyed Size	Medi Lution Health Care Beta Tech Turkey	Popular International Pvt. Ltd Parieten Macroporus Mesh , (Poly Propylene) , Size 15x15cm , Covidien Medtronic USA	J.J. Brothers Enterprises Duzey Medical			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Rejected Poor quality Material Difficult to handle	Accepted	Rejected Poor quality Material Difficult to handle			

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
206	Plasma Pheresis / Exchange Kits	The successful bidder is responsible to provide /Place/ install 1 machine at their own expenses (free of cost) in this hospital for the contract period (one year from issuance of AAT/notification of award). The successful bidder shall ensure that the machines remain functional / operational round the clock 24/7 to ensure that the patients are not refused due to non functioning of machine. The successful bidder is also responsible for repair / maintenance / upgradation and replacement of machine at their own expense as and when required. Similarly the successful bidder is also responsible to arrange the staff at their own expense to run the machine as well as to perform the procedure. The site for placement of the machine will be approved by the Authority.	Life – Tec Nigale						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
207	Disposable Shoe Cover Non-woven	Rate / No.	Ibrahim Enterprises CLASSIC	Zainabain Enterprises Faisalabad WELLMED	J.J. Brothers Enterprises				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Very poor quality	Accepted	Rejected Very poor quality				
208	Disposable Caps	Rate/No.	Ibrahim Enterprises BLUE BIRD	Zainabain Enterprises Faisalabad WELLMED	J.J. Brothers Enterprises				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Poor quality				
209	Surgical Face Mask with nose pin & ultrasonic punch with elastic strips	3-ply, Non-woven cloth size:6.90" - 7" X 3.5" - 3.75" Rate/No	Ibrahim Enterprises MEDX	Zainabain Enterprises Faisalabad WELLMED	Tanveer Enterprises Medpro Surgical Disposable Face Mask				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Central layer is poor material like tissue paper				

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
210	Mask with Reservoir bag	Rate/No	Ibrahim Enterprises HITECARE	J.J. Brothers Enterprises Well Lead					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
211	Hand Sanitizer	Alcohol 75% aqeous solution of Isopropyl Alcohol Alcohol (ethanol) (80%, volume / volume (v/v)) in an aqueous solution denatured; Bottle of 1 Liter with actuator assembly.	N/Q						
212	DCP Broad	Rate / No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		From 212 to 281 Due to single bidder participating in the tender, ortho implants will be re-advertised to enhance the competition among bidders hence pending						
213	DCP Narrow	Rate / No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
214	Small Fragment DCP	Rate / No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
215	K Wires	1.5mm, 2.0mm, 2.5mm, 3mm, 3.5mm	Suriquips	J.J. Brothers Enterprises Best quality					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
216	I/L Nail Femur	Rate / No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
217	I/L Nail Tibia	Rate / No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
218	Distal Tibial Lockij Plate	Rate / No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user.		Pending	Pending					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
219	Distal Femor Lockij Plate	Rate / No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
220	Broad & Narrow Lockij DCP	Full Range	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
221	Ilizarove Frame System	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
222	AMP	Rate / No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
223	Thompsor	Rate / No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
224	Bipolar Prosthesis	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
225	K.Nail	Rate/No	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
226	Drill bit	Full Range Rate/No	Suriquips	G.B. Trading Local	J.J. Brothers Enterprises Best quality				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending	Pending				

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
227	DHS with Barrel plate	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
228	DCS With Barrel Plate	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
229	Distal Radius TLCP	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
230	Cannulated Screw System	6.5mm	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
231	Cannulated Screw System	4mm	Suriquips	G.B. Trading Local					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
232	Circlage Wire	18 No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
233	Circlage Wire	20 No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
234	Cortical Screw 3.5mm Self tapping	Range 12-40	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
235	Cortical Screw 4.5mm Self tapping	Range 12-60	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user.		Pending	Pending					

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
236	1/3rd Tibular Plate 3.5mm	6-9 Range	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
237	L Butress Plate 4.5mm	R	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
238	L Butress Plate 4.5mm	L	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
239	T Butress Plate 4.5mm	R	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
240	T Butress Plate 4.5mm	L	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
241	Reconstruction Plates 3.5mm	Full Range	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
242	Reconstruction Plates 4.5mm	Full Range	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
243	Cancellous Screw 6.5mm	16 & 32 Threaded	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
244	Disposable C arm Cover	Rate/No	Suriquips	Tanveer Enterprises Medpro C Arm Cover					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
245	Washers	All sizes	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
246	Nail Guide Wire	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
247	DHS Guide Wire	Rate/No	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
248	Locking Screw	All sizes 5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
249	Cannulated Screw Driver	4mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
250	Cannulated Screw Driver	6.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
251	Cannulated Guide wire	4mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
252	Cannulated Guide wire	6.5mm	Suriquips						
	Product Evaluation by TEC / End user.		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
253	Locking Plate Instrument Set	4.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
254	Locking Plate Instrument Set	3.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
255	Proximal tibial anatomical locking plate	Right	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
256	Proximal tibial anatomical locking plate	Left	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user.		Pending	Pending					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
257	Recon Nail Femor		Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
258	Proximal Femor Nail		Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
259	PHILOS Plate	Rate/No.	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
260	Mini Fragment Plate	2.7mm	Suriquips						
	Product Evaluation by TEC / End user.		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
261	Rush Rod Radius	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
262	Rush Rod ulnar	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
263	Clavical S Plate	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
264	T Clamp For AO Fixator	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
265	Locking Drill Bit	4.5mm	Suriquips						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
266	Depth Gauge	3.5mm, 4.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
267	Drill seleve	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
268	Broken Set	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
269	Osteotome Set	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
270	TENS Nail	Rate/No.	Suriquips						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
271	Inter Locking Nail Screw Femor	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
272	Inter Locking Nail Screw Tibia	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
273	Cancellous Screw 4mm	16 Threaded , Size 14-16	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
274	Cancellous Screw 4mm	32 Threaded, Size 14-16	Suriquips	G.B. Trading Local					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending	Pending					
275	Total Hip Implant		N/Q						
276	Bone Cement		N/Q						
277	Hand Set	Rate/No.	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
278	Standard T-Plate 4.5mm	4.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
279	T-Plate 3.5mm	3.5mm	Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
280	Standard L Plate		Suriquips						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Pending						
281	Total Knee Replacement		N/Q						
282	Achalasia Balloon	30mm	Tahir Pharma						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Cook Accepted						
283	Achalasia Balloon	35mm	Tahir Pharma						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Cook Accepted						
284	Achalasia Balloon	40mm	Tahir Pharma						
	Product Evaluation by TEC / End user.		Marflow AG Rejected						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
285	Shooter	Rate/No.	Tahir Pharma Marflow AG						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
286	Sclerotherapy Inj. Needle	Rate/No.	Hope Enterprises Disposable Injector Needle	Tahir Pharma Marflow AG					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
287	Disposable Biospy Forcep	2.4mm, Length 160cm	Hope Enterprises Disposable Biopsy Forceps Oval Cup	Tahir Pharma Cook					
	Product Evaluation by TEC / End user.		Rejected Poor quality	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
288	Disposable Biospy Forcep	2.8mm, Length 240cm and Above	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
289	Esophageal Metallic Stent	18mm x10cm (Single USE)	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
290	Esophageal Metallic Stent	18mm x12cm (Single USE)	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
291	Esophageal Metallic Stent	18mm x15cm (Single USE)	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
292	Esophageal Metallic Stent	18mmx10.3cm (Single Use)	Tahir Pharma Marflow AG						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
293	Esophageal Metallic Stent	18mmx12/12.3cm (Single Use)	Tahir Pharma Marflow AG						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
294	Polypectomy Snare	3cm Loop Diameter 3cm, Working Length 240cm Or More Channel Diameter 2.8	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
295	Wire Guided CRE Baloon	CRE 8-9-10 mm Pebax Matrial, 3 Stage	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		Inflation Balloon (Single Use)							
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
296	Wire Guided CRE Baloon	CRE 10-11-12 mm Pebax Matrial, 3 Stage Inflation Balloon (Single Use)	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
297	Wire Guided CRE Baloon	CRE 12-13.5-125 mm Pebax Matrial, 3 Stage Inflation Balloon (Single Use)	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
298	Wire Guided CRE Baloon	CRE 15-16.5-18 mm (Single Use)	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
299	Wire Guided CRE Baloon	CRE 18-19-20 mm (Single Use)	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
300	Achalasia Inflation Device	Phematic Pump Quantum Inflation Device	Tahir Pharma Cook Accepted						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
301	Biliary Balloon	Quantum TTC Billiary Balloon Dilator 8mmx3cm	N/Q						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
302	Biliary Balloon	Quantum TTC Billiary Balloon Dilator 10mmx3/4cm	N/Q						
303	Biliary Cytology Brush	CytomaxII Double Lumen Cytology Brush 8Fr Catheter With 1.5 cm filiform ti 2.5cm length Brush single use)	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
304	Double Pitgail Biliary Stent	7Fr x 7cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
305	Double Pitgail Biliary Stent	10Fr x 3cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
306	ERCP Guide Wire	Tracer Metro Direct Wire Guide	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		0.035x450cm And Above							
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
307	ERCP Guide Wire	Tracer Metro Direct Wire Guide / 0.025x450cm And Above	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
308	Hemostatic Clip Lower	Instinct Endoscopic Hemoclip (16mm)	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
309	Hemostatic Clip Uper	Instinct Endoscopic Hemoclip (16mm)	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
310	Needle Knife	Huibregtse-Triple Lumen Needle Knife 7Fr/200cm pre Curved catheter tapered to 5Fr , Adjustable 4mm Cutting Wire	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
311	Pancreatic plastic Stent (Single Pigtail)	Zimmon Pancreatic Stent 5 Fr x 5cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
312	Pancreatic plastic Stent (Single Pigatail)	Zimmon Pancreatic Stent 7 Fr x 7cm	Tahir Pharma						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Cook Accepted						
313	Papillotome	TRI-30 TRI-Tome PC Triple Lumer Sphincterotome 30mm Cut Wire , Tip Length 5, Shaft O.D 5.5F (Single Use)	Tahir Pharma						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Cook Accepted						
314	Soehinder Biliary Dilator	Soehendra Biliary Dilation Catheter 8.5 Fr 200cm length	Tahir Pharma						
	Product Evaluation by TEC / End user.		Cook Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
315	Soehinder Biliary Dilator	10 Fr 200cm length	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
316	Stent Pusher	Pushing Catheter 7 Fr	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
317	Stone Extraction Ballon	Fusion Quattro Extraction Balloon-12-15-18-20-A	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
318	Straight Biliary Plastic Stent	Cotton Leung Biliary Stent 7FrX7cm	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
319	Straight Biliary Plastic Stent	Cotton Leung Biliary Stent 7FrX10cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
320	Straight Biliary Plastic Stent	Cotton Leung Biliary Stent 10FrX10cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
321	Straight Biliary Plastic Stent	Cotton Leung Biliary Stent 7FrX12cm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
322	Rocker Switch Pencil Reuseable	Diathermy Lead Rate/No.	Popular International Pvt. Ltd Reusable Rocker Switch Pencil, Covidien Medtronic USA						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
323	Rigid Lens	6.5mm	N/Q						
324	Opsite with Bag	200mmx200mm	N/Q						
325	Gold Probe	Gold Probe7f	N/Q						
326	Gold Probe	Gold Probe10f	N/Q						
327	Biliary Metallic Stent (Partially Covered)	10mm*80mm	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Re-tender						
328	Biliary Metallic Stent (Partially Covered)	10mm*100mm	Tahir Pharma Cook						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
329	Hemospary	Hemospray Endoscopic Hempstat 10fr/220cm Catheter (Single Use)	N/Q						
330	Peg Tube	PEG-24-Pull Method-1-S Percutaneos Endoscopic Gastrostomy Set 24 FR	J.J. Brothers Enterprises Well Lead	Tahir Pharma Cook					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Item pending						
331	Nephrostomy set	7Fr – 8.5Fr	J.J. Brothers Enterprises SCW Medcath China	Tahir Pharma Marflow AG					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Not according to specification provided PCN tube instead of nephrostomy set	Accepted					
332	Roth net for foreign body		N/Q						
333	Four prong foreign body remover		N/Q						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
334	Alligator forceps		Hope Enterprises Disposable Biopsy Forceps Alligator Jaw						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
335	Rat tooth forceps		Hope Enterprises Disposable Grasping Forceps Rat tooth						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
336	Stone crusher basket		Hope Enterprises						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
			Disposable Stone Retrieval basket						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
337	Stone crusher handle		N/Q						
338	Sphincteroplasty balloon wire guided	3cm x 12-15mm	N/Q						
339	Haemoclip assembly		N/Q						
340	Haemoclips for assembly		N/Q						
341	Stent Pusher assembly	Pushing Catheter 10 Fr	Tahir Pharma Cook						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
342	Tight Rope Button	For ACL	N/Q						
343	Bioabsorbable Screw	For ACL	N/Q						
344	Shevers	For ACL	N/Q						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
345	Left sided double lumen endobronchial tube	Size 35,37,39	Popular International Pvt. Ltd Shiley Covidien Medtronic USA	J.J. Brothers Enterprises Well Lead					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
346	Right sided double lumen endobronchial tube	Size 35,37,39	Popular International Pvt. Ltd Shiley Covidien Medtronic USA	J.J. Brothers Enterprises Well Lead					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
347	Blood Leucocytes Reduction Filter Bed side (Zisc Technology/FDA/NANDO Approved)		G.B. Trading PuriBlood Taiwan Accepted	Tahir Pharma Leukocyte Filter					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected			Rejected Poor packing quality Does not advance zisc technology					
348	Chlorhexidine Gauze dressing	15cm*1.5M Rate/Roll	Friends Enterprises PARA-TULLE	Zahid Traders Septa Chlor Faisal Pharma	Tanveer Enterprises Septi Grass				
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Quality not upto the mark	Accepted	Rejected Quality not upto the mark				
349	Polyglactin / Poly glycolic Acid Braided, Synthetic, Absorbable suture No. 5/0	Round bodied, with needle, Rate/No.	Muller & Phipps Pakistan Sutures Absorbable Vicryl Johnson & Johnson						
	Product Evaluation by TEC / End user.		Accepted						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
350	Polyglactin / Poly glycolic Acid Braided, Synthetic, Absorbable suture No. 6/0	Round bodied, with needle, Rate/No.	N/Q						
351	Amplatz Sheath	All Sizes	N/Q						
352	Carbiod Burr Length 70mm For (ENT)	Head Size 2mm, 3mm, 4mm	G.B. Trading Local						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Sample not provided						
353	Poly Propylene 7/0	Round bodied, with needle, Rate/No.	Muller & Phipps Pakistan Sutures None Absorb Prolene Johnson & Johnson						
	Product Evaluation by TEC / End user.		Sample not provided						

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
354	Black Silk No. 3/0	Round bodied, with needle, Rate/No.	Ibrahim Enterprises B.T. SILK SUTURE	Muller & Phipps Pakistan Sutures None Absorb Silk Johnson & Johnson					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Rejected Poor quality Poor strength	Accepted					
355	Arterial Line with interated Guide wire	(Adult/20g)	Tanveer Enterprises ARROW	Tahir Pharma Intra Special Catheter					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Not according to specification Without integrated wire					
356	Arterial Line with interated Guide wire	(Paeds/22G)	Tanveer Enterprises ARROW	Tahir Pharma					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
				Intra Special Catheter					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Not according to specification Without integrated wire					
357	CVP Line	(Paeds)	Friends Enterprises Fornia	Flowtronix Systems Able	Popular International Pvt. Ltd GM-Key Giant Medical Ltd	Tanveer Enterprises ARROW			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted	Rejected Not according to specification Provided adult guage for paeds CVP	Accepted			
358	Bare Laser Emission Fibre 600µm	Bare Laser Emission Fibre 600µm, Standard SMA 905 connector with knurl nut, Capillary OD 950µm, NA 0.37, length 3m for hemorrhoid, compatible with Generator. (CE, ISO, FDA) Generator will be (F.O.C.)	Popular International Pvt. Ltd Bare Fiber 600µm For Hemorrhoids Gigaa Laser Technology Inc. China	Tahir Pharma Bioletic					

Pharmacist, Purchase Cell

Director Purchase FMU

Dr. Bashir Ahmed Gondal
Assistant Professor of Surgery

Dr. Muhammad Saleem Iqbal
Assistant Professor of Surgery

Dr. Sajid Rehman Randhawa
Associate Professor of Surgery

Prof. Dr. Muhammad Faisal Bilal Lodhi
Head of Surgical Unit-II, AHF-I

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
359	Bare Laser Emission Fibre 400µm	Bare Laser Emission Fibre 400µm, Standard SMA 905 connector with knurl nut, Capillary OD 950µm, NA 0.37, length 3m for hemorrhoid, compatible with Generator. (CE, ISO, FDA) Generator will be (F.O.C.)	Popular International Pvt. Ltd Bare Fiber 400µm For Hemorrhoids Gigaa Laser Technology Inc. China	Tahir Pharma Bioletic					
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
360	Radial Emission Fibre 600µm	White Tefzel Jacket with length marking, standard SMA 905 connector with knurl nut, PVC protection cap, extension sleeve red engraved, black bend protection, distal conical tip with silica capillary dome, capillary OD 1800µm,	Popular International Pvt. Ltd Radia Fiber 600µm for Varicose Vein Gigaa Laser Technology Inc. China	Tahir Pharma Bioletic					

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
		NA 0.22, length 2.6m for varicose vein, compatible with Generator. (CE, ISO, FDA) Generator will be (F.O.C.)							
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted	Accepted					
361	Disposable Grasping Forceps 3 PRONG HOOK TYPE Gastroscopy	Diameter 2.4mm, Length 1600mm Channel 2.8mm	Hope Enterprises Disposable Grasping Forceps 3 PRONG						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
362	Disposable Grasping Forceps 5 PRONG HOOK TYPE Gastroscopy	Diameter 2.4mm, Length 1800mm Channel 2.8mm	Hope Enterprises Disposable Grasping Forceps 5PRONG						
	Product Evaluation by TEC / End user.		Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
363	Disposable Grasping Forceps Net Type Gastroscopy	Diameter 2.4mm, Length 1800mm Channel 2.8mm	Hope Enterprises Disposable Grasping Forceps Rothnet Type						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
364	Disposable Biopsy Forceps Oval Cup without Needle	Diameter 2.4mm, Length 2300mm	Hope Enterprises Disposable Biopsy Forceps Oval Cup						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
365	Disposable Biopsy Forceps Oval Cup with Needle	Diameter 2.4mm, Length 2300mm	Hope Enterprises Disposable Biopsy Forceps Oval Cup						
	Product Evaluation by TEC / End user.		Accepted						

EVALUATION OF SURGICAL & DISPOSABLE ITEMS FMU, ALLIED HOSPITAL-I & ALLIED HOSPITAL-II, FTH FAISALABAD 2024-2025

Tender item #	NAME OF ITEMS	SPECIFICATIONS	Name of firms						
	Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected								
366	Disposable Biopsy Forceps Alligator Jaw without Needle	Diameter 2.4mm, Length 2300mm	Hope Enterprises Disposable Biopsy Forceps Alligator Jaw						
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Accepted						
367	Flow Regulator	2 to 400ml/h with y port	Friends Enterprises ZQR	Ibrahim Enterprises NATIONAL	J.J. Brothers Enterprises Golden	Tahir Pharma B-Flow			
	Product Evaluation by TEC / End user. Product reported of poor quality on the base of clinical data / Past experience / sample evaluation will be rejected		Not according to specification	Not according to specification	Not according to specification	Accepted			